




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020217\
Data File : BF092761.D
Acq On : 3 Feb 2017 4:34
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 03 07:05:43 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 30 14:50:05 2017
Response via : Initial Calibration



TIC: BF092761.D

(69) Pentachlorophenol (C)




11.265min (-0.034) 57756.30ng

response 92338

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	63.73
264.00	62.90	62.22
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020217\
Data File : BF092761.D
Acq On : 3 Feb 2017 4:34
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 03 07:05:43 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 30 14:50:05 2017
Response via : Initial Calibration



TIC: BF092761.D

(76) Benzidine

12.796min (-0.034) 0.00ng

response 744831

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	14.39
183.00	11.50	11.76
0.00	0.00	0.00