

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020222\
 Data File : BF126785.D
 Acq On : 02 Feb 2022 14:21
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICBF020222

Quant Time: Feb 02 16:46:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 15:08:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.739	0.725	1.9	100	0.00
3	Pyridine	2.004	2.011	-0.3	101	0.00
4	n-Nitrosodimethylamine	0.619	0.610	1.5	99	0.00
5 S	2-Fluorophenol	1.460	1.420	2.7	103	0.00
6	Aniline	2.577	2.541	1.4	99	0.00
7 S	Phenol-d6	2.020	1.962	2.9	101	0.00
8	2-Chlorophenol	1.335	1.301	2.5	101	0.00
9	Benzaldehyde	1.271	1.226	3.5	102	0.00
10 C	Phenol	2.162	2.085	3.6	101	0.00
11	bis(2-Chloroethyl)ether	1.742	1.691	2.9	101	0.00
12	1,3-Dichlorobenzene	1.503	1.455	3.2	101	0.00
13 C	1,4-Dichlorobenzene	1.520	1.467	3.5	100	0.00
14	1,2-Dichlorobenzene	1.477	1.369	7.3	101	0.00
15	Benzyl Alcohol	1.614	1.580	2.1	99	0.00
16	2,2'-oxybis(1-Chloropropane	1.876	1.840	1.9	99	0.00
17	2-Methylphenol	1.313	1.281	2.4	100	0.00
18	Hexachloroethane	0.595	0.568	4.5	99	0.00
19 P	n-Nitroso-di-n-propylamine	1.330	1.259	5.3	99	0.00
20	3+4-Methylphenols	1.654	1.608	2.8	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.579	0.558	3.6	101	0.00
23 S	Nitrobenzene-d5	0.534	0.523	2.1	99	0.00
24	Nitrobenzene	0.530	0.517	2.5	99	0.00
25	Isophorone	0.917	0.889	3.1	99	0.00
26 C	2-Nitrophenol	0.177	0.182	-2.8	100	0.00
27	2,4-Dimethylphenol	0.309	0.306	1.0	100	0.00
28	bis(2-Chloroethoxy)methane	0.582	0.567	2.6	98	0.00
29 C	2,4-Dichlorophenol	0.304	0.298	2.0	98	0.00
30	1,2,4-Trichlorobenzene	0.327	0.310	5.2	97	0.00
31	Naphthalene	1.045	1.013	3.1	99	0.00
32	Benzoic acid	0.238	0.242	-1.7	96	0.00
33	4-Chloroaniline	0.422	0.412	2.4	99	0.00
34 C	Hexachlorobutadiene	0.205	0.196	4.4	99	0.00
35	Caprolactam	0.111	0.111	0.0	98	0.00
36 C	4-Chloro-3-methylphenol	0.387	0.376	2.8	99	0.00
37	2-Methylnaphthalene	0.644	0.627	2.6	99	0.00
38	1-Methylnaphthalene	0.620	0.611	1.5	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	0.576	0.550	4.5	99	0.00
41 P	Hexachlorocyclopentadiene	0.332	0.320	3.6	96	0.00
42 S	2,4,6-Tribromophenol	0.208	0.201	3.4	99	0.00
43 C	2,4,6-Trichlorophenol	0.406	0.385	5.2	98	0.00
44	2,4,5-Trichlorophenol	0.424	0.414	2.4	100	0.00
45 S	2-Fluorobiphenyl	1.208	1.167	3.4	99	0.00
46	1,1'-Biphenyl	1.465	1.392	5.0	99	0.00
47	2-Chloronaphthalene	1.161	1.113	4.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020222\
 Data File : BF126785.D
 Acq On : 02 Feb 2022 14:21
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF020222

Quant Time: Feb 02 16:46:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 15:08:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.473	0.467	1.3	100	0.00
49	Acenaphthylene	1.793	1.720	4.1	99	0.00
50	Dimethylphthalate	1.382	1.330	3.8	100	0.00
51	2,6-Dinitrotoluene	0.284	0.280	1.4	100	0.00
52 C	Acenaphthene	1.143	1.107	3.1	100	0.00
53	3-Nitroaniline	0.324	0.321	0.9	99	0.00
54 P	2,4-Dinitrophenol	0.150	0.159	-6.0	101	0.00
55	Dibenzofuran	1.687	1.582	6.2	98	0.00
56 P	4-Nitrophenol	0.245	0.246	-0.4	101	0.00
57	2,4-Dinitrotoluene	0.385	0.379	1.6	100	0.00
58	Fluorene	1.248	1.189	4.7	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.367	0.355	3.3	100	0.00
60	Diethylphthalate	1.358	1.278	5.9	98	0.00
61	4-Chlorophenyl-phenylether	0.631	0.593	6.0	98	0.00
62	4-Nitroaniline	0.307	0.304	1.0	97	0.00
63	Azobenzene	1.885	1.815	3.7	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	0.121	0.135	-11.6	101	0.00
66 c	n-Nitrosodiphenylamine	0.643	0.640	0.5	101	0.00
67	4-Bromophenyl-phenylether	0.229	0.228	0.4	98	0.00
68	Hexachlorobenzene	0.250	0.246	1.6	100	0.00
69	Atrazine	0.209	0.211	-1.0	100	0.00
70 C	Pentachlorophenol	0.146	0.154	-5.5	101	0.00
71	Phenanthrene	1.108	1.066	3.8	98	0.00
72	Anthracene	1.109	1.074	3.2	100	0.00
73	Carbazole	1.047	0.996	4.9	99	0.00
74	Di-n-butylphthalate	1.213	1.178	2.9	98	0.00
75 C	Fluoranthene	1.127	1.082	4.0	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzydine	0.459	0.491	-7.0	100	0.00
78	Pyrene	1.598	1.542	3.5	100	0.00
79 S	Terphenyl-d14	1.196	1.119	6.4	99	0.00
80	Butylbenzylphthalate	0.663	0.663	0.0	101	0.00
81	Benzo(a)anthracene	1.335	1.306	2.2	104	0.00
82	3,3'-Dichlorobenzidine	0.382	0.375	1.8	102	0.00
83	Chrysene	1.264	1.236	2.2	105	0.00
84	Bis(2-ethylhexyl)phthalate	0.863	0.860	0.3	103	0.00
85 c	Di-n-octyl phthalate	1.230	1.223	0.6	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	1.220	1.239	-1.6	97	0.00
88	Benzo(b)fluoranthene	1.308	1.310	-0.2	106	0.00
89	Benzo(k)fluoranthene	1.305	1.320	-1.1	104	0.00
90 C	Benzo(a)pyrene	1.042	1.036	0.6	103	0.00
91	Dibenzo(a,h)anthracene	0.999	1.023	-2.4	98	0.00
92	Benzo(g,h,i)perylene	0.987	1.031	-4.5	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020222\
Data File : BF126785.D
Acq On : 02 Feb 2022 14:21
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF020222

Quant Time: Feb 02 16:46:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 02 15:08:47 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0