

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020315\
 Data File : BF077170.D
 Acq On : 3 Feb 2015 20:01
 Operator : TP/IZ
 Sample : G1199-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-051-13015

Quant Time: Feb 04 02:22:45 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 01:28:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	24243	20.00	ng	0.00
21) Naphthalene-d8	10.59	136	96129	20.00	ng	0.00
38) Acenaphthene-d10	14.71	164	44888	20.00	ng	-0.01
63) Phenanthrene-d10	17.56	188	85349	20.00	ng	0.00
75) Chrysene-d12	21.07	240	83596	20.00	ng	0.00
86) Perylene-d12	22.73	264	79177	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	4.70	112	63149	42.83	ng	0.01
7) Phenol-d6	7.10	99	44780	24.07	ng	0.01
23) Nitrobenzene-d5	8.99	82	151667	87.41	ng	0.00
41) 2,4,6-Tribromophenol	16.44	330	51797	142.16	ng	0.00
44) 2-Fluorobiphenyl	13.24	172	301864	102.34	ng	0.00
78) Terphenyl-d14	19.81	244	284258	78.28	ng	0.00

Target Compounds						Qvalue
3) Pyridine	1.61	79	11380	6.97	ng	# 88
6) Aniline	6.98	93	21738	8.44	ng	98
13) 1,4-Dichlorobenzene	7.72	146	27544	13.43	ng	100
14) 1,2-Dichlorobenzene	8.04	146	477283	252.58	ng	98
30) 1,2,4-Trichlorobenzene	10.51	180	186119	131.53	ng	93
31) Naphthalene	10.64	128	227147	45.90	ng	99
33) 4-Chloroaniline	10.90	127	97696	48.85	ng	95
37) 2-Methylnaphthalene	12.29	142	51936	15.37	ng	94
45) 1,1'-Biphenyl	13.45	154	9939	2.99	ng	88
49) Dimethylphthalate	14.32	163	11151	3.50	ng	# 95
51) Acenaphthene	14.79	154	14538	5.55	ng	97
54) Dibenzofuran	15.22	168	10343	2.71	ng	# 97
56) 2,4-Dinitrotoluene	15.36	165	17435	19.59	ng	# 87
57) Fluorene	15.97	166	18433	5.70	ng	98
70) Phenanthrene	17.60	178	13124	2.47	ng	96
72) Carbazole	17.97	167	22013	4.37	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020315\
Data File : BF077170.D
Acq On : 3 Feb 2015 20:01
Operator : TP/IZ
Sample : G1199-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-051-13015

Quant Time: Feb 04 02:22:45 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 04 01:28:22 2015
Response via : Initial Calibration

