

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020316\
 Data File : BF084786.D
 Acq On : 3 Feb 2016 13:28
 Operator : SJ/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 04 00:37:09 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 14:16:01 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.562	0.539	4.1	96	0.00
3	Pyridine	1.465	1.589	-8.5	112	0.00
4	n-Nitrosodimethylamine	0.758	0.730	3.7	93	0.00
5 S	2-Fluorophenol	1.284	1.355	-5.5	110	0.00
6	Aniline	1.948	1.958	-0.5	100	0.00
7 S	Phenol-d6	1.592	1.540	3.3	101	0.00
8	2-Chlorophenol	1.453	1.546	-6.4	103	0.00
9	Benzaldehyde	0.940	0.942	-0.2	96	0.00
10 C	Phenol	1.783	1.608	9.8	101	0.00
11	bis(2-Chloroethyl)ether	1.391	1.321	5.0	101	0.00
12	1,3-Dichlorobenzene	1.536	1.523	0.8	102	0.00
13 C	1,4-Dichlorobenzene	1.535	1.562	-1.8	104	0.00
14	1,2-Dichlorobenzene	1.430	1.373	4.0	91	0.00
15	Benzyl Alcohol	1.099	1.069	2.7	98	0.00
16	2,2'-oxybis(1-Chloropropane	2.339	2.218	5.2	98	0.00
17	2-Methylphenol	1.149	1.180	-2.7	96	0.00
18	Hexachloroethane	0.591	0.579	2.0	95	0.00
19 P	n-Nitroso-di-n-propylamine	0.998	0.938	6.0	98	0.00
20	3+4-Methylphenols	1.427	1.467	-2.8	102	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.527	0.532	-0.9	101	0.00
23 S	Nitrobenzene-d5	0.355	0.367	-3.4	99	0.00
24	Nitrobenzene	0.380	0.351	7.6	100	0.00
25	Isophorone	0.690	0.693	-0.4	97	0.00
26 C	2-Nitrophenol	0.188	0.200	-6.4	105	0.00
27	2,4-Dimethylphenol	0.326	0.319	2.1	102	0.00
28	bis(2-Chloroethoxy)methane	0.410	0.413	-0.7	102	0.00
29 C	2,4-Dichlorophenol	0.279	0.298	-6.8	100	0.00
30	1,2,4-Trichlorobenzene	0.315	0.311	1.3	99	0.00
31	Naphthalene	1.003	0.940	6.3	98	0.00
32	Benzoic acid	0.209	0.229	-9.6	107	0.00
33	4-Chloroaniline	0.415	0.394	5.1	102	0.00
34 C	Hexachlorobutadiene	0.182	0.192	-5.5	102	0.00
35	Caprolactam	0.086	0.083	3.5	83	0.00
36 C	4-Chloro-3-methylphenol	0.318	0.317	0.3	104	0.00
37	2-Methylnaphthalene	0.628	0.671	-6.8	99	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
39	1,2,4,5-Tetrachlorobenzene	0.635	0.589	7.2	95	0.00
40 P	Hexachlorocyclopentadiene	0.223	0.252	-13.0	105	0.00
41 S	2,4,6-Tribromophenol	0.209	0.226	-8.1	96	0.00
42 C	2,4,6-Trichlorophenol	0.409	0.418	-2.2	94	0.00
43	2,4,5-Trichlorophenol	0.416	0.409	1.7	96	0.00
44 S	2-Fluorobiphenyl	1.321	1.283	2.9	102	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020316\
 Data File : BF084786.D
 Acq On : 3 Feb 2016 13:28
 Operator : SJ/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 04 00:37:09 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 14:16:01 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.786	1.819	-1.8	95	0.00
46	2-Chloronaphthalene	1.316	1.252	4.9	91	0.00
47	2-Nitroaniline	0.417	0.378	9.4	96	0.00
48	Acenaphthylene	1.996	1.929	3.4	90	0.00
49	Dimethylphthalate	1.523	1.605	-5.4	98	0.00
50	2,6-Dinitrotoluene	0.333	0.352	-5.7	94	0.00
51 C	Acenaphthene	1.299	1.229	5.4	89	0.00
52	3-Nitroaniline	0.374	0.367	1.9	95	0.00
53 P	2,4-Dinitrophenol	0.133	0.150	-12.8	99	0.00
54	Dibenzofuran	1.587	1.586	0.1	93	0.00
55 P	4-Nitrophenol	0.275	0.284	-3.3	88	0.00
56	2,4-Dinitrotoluene	0.424	0.434	-2.4	94	0.00
57	Fluorene	1.326	1.239	6.6	86	0.00
58	2,3,4,6-Tetrachlorophenol	0.317	0.310	2.2	103	0.00
59	Diethylphthalate	1.487	1.479	0.5	96	0.00
60	4-Chlorophenyl-phenylether	0.621	0.600	3.4	97	0.00
61	4-Nitroaniline	0.364	0.332	8.8	88	0.00
62	Azobenzene	1.430	1.452	-1.5	97	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	0.123	0.127	-3.3	96	0.00
65 c	n-Nitrosodiphenylamine	0.635	0.656	-3.3	95	0.00
66	4-Bromophenyl-phenylether	0.221	0.225	-1.8	94	0.00
67	Hexachlorobenzene	0.241	0.222	7.9	96	0.00
68	Atrazine	0.201	0.175	12.9	92	0.00
69 C	Pentachlorophenol	0.113	0.114	-0.9	95	0.00
70	Phenanthrene	1.109	1.020	8.0	97	0.00
71	Anthracene	1.118	1.122	-0.4	93	0.00
72	Carbazole	0.975	0.949	2.7	89	0.00
73	Di-n-butylphthalate	1.241	1.108	10.7	92	0.00
74 C	Fluoranthene	1.123	1.070	4.7	90	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
76	Benzidine	0.597	0.624	-4.5	79	0.00
77	Pyrene	1.531	1.475	3.7	84	0.00
78 S	Terphenyl-d14	0.883	0.961	-8.8	103	0.00
79	Butylbenzylphthalate	0.695	0.711	-2.3	87	0.00
80	Benzo(a)anthracene	1.241	1.212	2.3	88	0.00
81	3,3'-Dichlorobenzidine	0.440	0.413	6.1	77	0.00
82	Chrysene	1.204	1.112	7.6	81	0.00
83	Bis(2-ethylhexyl)phthalate	0.864	0.885	-2.4	90	0.00
84 c	Di-n-octyl phthalate	1.426	1.312	8.0	81	0.00
85	Indeno(1,2,3-cd)pyrene	0.947	0.930	1.8	90	0.00
86 I	Perylene-d12	1.000	1.000	0.0	81	0.00
87	Benzo(b)fluoranthene	1.344	1.222	9.1	73	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020316\
Data File : BF084786.D
Acq On : 3 Feb 2016 13:28
Operator : SJ/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 04 00:37:09 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 03 14:16:01 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.111	1.257	-13.1	94	0.00
89 C	Benzo(a)pyrene	1.145	1.165	-1.7	82	0.00
90	Dibenzo(a,h)anthracene	0.964	1.050	-8.9	90	0.00
91	Benzo(a,h,i)perylene	0.971	1.097	-13.0	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0