


Data Path : Z:\HPCHEM1\BNA F\DATA\BF020316\
 Data File : BF084785.D
 Acq On : 3 Feb 2016 11:54
 Operator : SJ/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 03 14:16:01 2016



AutoFind: Scans 821, 822, 823; Background Corrected with Scan 817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	24349	PASS
68	69	0.00	2	2.1	450	PASS
69	198	0.00	100	33.2	21460	PASS
70	69	0.00	2	0.8	172	PASS
127	198	10	80	47.9	30992	PASS
197	198	0.00	2	0.7	447	PASS
198	198	100	100	100.0	64733	PASS
199	198	5	9	6.8	4373	PASS
275	198	10	60	24.8	16054	PASS
365	198	1	100	3.0	1970	PASS
441	442	0.01	24	16.7	9234	PASS
442	198	50	100	85.3	55222	PASS
443	442	15	24	19.3	10656	PASS