

Data Path : U:\HPCHEM1\BNA F\DATA\BF020318\
 Data File : BF102717.D
 Acq On : 3 Feb 2018 15:21
 Operator : SJ/JU
 Sample : J1410-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Feb 05 01:23:47 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 00:38:27 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	194623	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	800070	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	303758	20.00	ng	0.00
63) Phenanthrene-d10	11.42	188	281168	20.00	ng	0.02
75) Chrysene-d12	14.04	240	369742	20.00	ng	0.00
86) Perylene-d12	15.51	264	395444	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	1313351	102.48	ng	0.00
7) Phenol-d6	6.50	99	1554892	100.77	ng	0.00
23) Nitrobenzene-d5	7.44	82	968866	84.01	ng	0.00
41) 2,4,6-Tribromophenol	10.71	330	235840	96.46	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	1460021	79.76	ng	0.00
78) Terphenyl-d14	13.00	244	881703	56.46	ng	0.01
Target Compounds						
6) Aniline	6.52	93	277775	12.190	ng	# 50
9) Benzaldehyde	6.52	77	114283	9.973	ng	# 5
10) Phenol	6.50	94	77962	4.714	ng	# 59
11) bis(2-Chloroethyl)ether	6.63	93	91030	6.615	ng	# 23
19) n-Nitroso-di-n-propylamine	7.44	70	132871	13.060	ng	# 82
24) Nitrobenzene	7.49	77	115	Below Cal		# 38
25) Isophorone	7.44	82	968749	36.478	ng	# 64
26) 2-Nitrophenol	7.87	139	125	8.685	ng	# 1
32) Benzoic acid	7.94	122	113	8.557	ng	# 29
47) 2-Nitroaniline	9.49	65	6492	4.441	ng	# 52
49) Dimethylphthalate	9.63	163	122051	5.153	ng	# 90
53) 2,4-Dinitrophenol	10.11	184	479	8.530	ng	# 1
55) 4-Nitrophenol	10.02	139	4444	4.950	ng	# 1
56) 2,4-Dinitrotoluene	10.07	165	4691	3.802	ng	# 35
64) 4,6-Dinitro-2-methylphenol	10.53	198	953	9.545	ng	# 1
68) Atrazine	11.11	200	9506	3.249	ng	# 1
74) Fluoranthene	12.63	202	39409	2.501	ng	# 77
79) Butylbenzylphthalate	13.46	149	14513	2.049	ng	# 86
83) Bis(2-ethylhexyl)phthalate	14.02	149	40612	2.287	ng	# 92
87) Benzo(b)fluoranthene	15.07	252	72464	2.826	ng	# 87
88) Benzo(k)fluoranthene	15.07	252	72464	2.958	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF020318\
Data File : BF102717.D
Acq On : 3 Feb 2018 15:21
Operator : SJ/JU
Sample : J1410-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Feb 05 01:23:47 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 05 00:38:27 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.87 min Scan# 814
Delta R.T. -0.01 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 194623
Ion Ratio Lower Upper
152 100
150 161.0 124.6 186.8
115 65.9 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 102.482 ng
RT: 5.49 min Scan# 579
Delta R.T. 0.01 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21

Tgt Ion:112 Resp: 1313351
Ion Ratio Lower Upper
112 100
64 63.5 47.9 71.9
63 32.1 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#6

Aniline

Concen: 12.190 ng

RT: 6.52 min Scan# 754

Delta R.T. -0.02 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 93 Resp: 277775

Ion Ratio Lower Upper

93 100

66 2.0 32.2 48.2#




65 7.7 16.4 24.6#



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1



#7

Phenol-d6

Concen: 100.767 ng

RT: 6.50 min Scan# 750

Delta R.T. 0.00 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion: 99 Resp: 1554892

Ion Ratio Lower Upper

99 100

42 20.9 14.5 21.7



71 33.9 25.4 38.0




Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1





#9

Benzaldehyde

Concen: 9.973 ng

RT: 6.52 min Scan# 754

Delta R.T. 0.09 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 77 Resp: 114283

Ion Ratio Lower Upper

77 100

105 6.5 81.1 121.1#




106 1.6 74.5 114.5#



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 4.714 ng

RT: 6.50 min Scan# 751

Delta R.T. -0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion: 94 Resp: 77962

Ion Ratio Lower Upper

94 100

65 39.3 7.9 47.9



66 69.4 16.2 56.2#




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1





#11

bis(2-Chloroethyl)ether

Concen: 6.615 ng

RT: 6.63 min Scan# 772

Delta R.T. 0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 93 Resp: 91030

Ion Ratio Lower Upper

93 100

63 3.8 58.6 98.6#

95 1.4 11.8 51.8#



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



Ion 95.00 (94.70 to 95.70): BF1

6.63

100000

50000

Time-->



#19

n-Nitroso-di-n-propylamine

Concen: 13.060 ng

RT: 7.44 min Scan# 910

Delta R.T. 0.15 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Tgt Ion: 70 Resp: 132871


Ion Ratio Lower Upper

70 100

42 53.2 47.5 71.3

101 0.0 7.4 11.2#

130 2.0 16.6 25.0#



Abundance Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): BF1

Ion 130.00 (129.70 to 130.70): BF1

7.44


200000


150000

100000

50000

Time-->





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.16 min Scan# 1032

Delta R.T. -0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 800070


Ion Ratio Lower Upper

136 100

137 11.2 8.9 13.3

54 10.3 6.6 9.8#

68 6.8 5.0 7.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#23

Nitrobenzene-d5

Concen: 84.006 ng

RT: 7.44 min Scan# 910

Delta R.T. -0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion: 82 Resp: 968866

Ion Ratio Lower Upper

82 100

128 46.7 36.1 54.1



54 61.9 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1





#25

Isophorone

Concen: 36.478 ng

RT: 7.44 min Scan# 910

Delta R.T. -0.26 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :
BNA_F
ClientSampleId :


Tot Ion: 82 Resp: 968749

Ion Ratio Lower Upper




82 100

95 0.1 5.2 7.8#

138 0.0 14.9 22.3#



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 8.685 ng

RT: 7.87 min Scan# 984

Delta R.T. 0.10 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion:139 Resp: 125

Ion Ratio Lower Upper



139 100


109 366.9 22.7 34.1#

65 0.0 40.5 60.7#



Abundance Ion 139.00 (138.70 to 139.70): BF1
Ion 109.00 (108.70 to 109.70): BF1
Ion 65.00 (64.70 to 65.70): BF1









#32
Benzoic acid
Concen: 8.557 ng
RT: 7.94 min Scan# 995
Delta R.T. 0.02 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21




Instrument :
BNA_F
ClientSampled :


Tot Ion:122 Resp: 113
Ion Ratio Lower Upper
122 100
105 0.0 101.1 141.1#
77 110.7 70.7 110.7



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.92 min Scan# 1331
Delta R.T. 0.00 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21

Tgt Ion:164 Resp: 303758
Ion Ratio Lower Upper
164 100
162 99.7 86.1 129.1
160 43.8 38.3 57.5









#41
 2,4,6-Tribromophenol
 Concen: 96.462 ng
 RT: 10.71 min Scan# 1466
 Delta R.T. 0.01 min
 Lab File: BF102717.D
 Acq: 3 Feb 2018 15:21




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:330 Resp: 235840
 Ion Ratio Lower Upper
 330 100
 332 101.6 77.0 115.6
 141 37.8 29.9 44.9



#44
 2-Fluorobiphenyl
 Concen: 79.757 ng
 RT: 9.23 min Scan# 1215
 Delta R.T. -0.01 min
 Lab File: BF102717.D
 Acq: 3 Feb 2018 15:21

Tgt Ion:172 Resp: 1460021
 Ion Ratio Lower Upper
 172 100
 171 36.2 29.7 44.5
 170 24.1 19.6 29.4





#47

2-Nitroaniline

Concen: 4.441 ng

RT: 9.49 min Scan# 1258

Delta R.T. 0.03 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampled :





Tot Ion: 65 Resp: 6492

Ion Ratio Lower Upper

65 100

92 43.1 55.8 83.6#

138 52.7 91.2 136.8#



#49

Dimethylphthalate

Concen: 5.153 ng

RT: 9.63 min Scan# 1282

Delta R.T. -0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21




Tgt Ion: 163 Resp: 122051


Ion Ratio Lower Upper

163 100

194 12.7 2.7 4.1#

164 11.6 8.2 12.2





#53

2,4-Dinitrophenol

Concen: 8.530 ng

RT: 10.11 min Scan# 1364

Delta R.T. 0.15 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampled :


Tot Ion:184 Resp: 479

Ion Ratio Lower Upper

184 100

63 528.1 54.2 81.2#




154 668.3 184.0 276.0#



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 154.00 (153.70 to 154.70): E



#55

4-Nitrophenol

Concen: 4.950 ng

RT: 10.02 min Scan# 1348

Delta R.T. 0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion:139 Resp: 4444

Ion Ratio Lower Upper

139 100

109 894.8 48.4 88.4#



65 142.3 72.6 112.6#




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1









#56
2,4-Dinitrotoluene
Concen: 3.802 ng
RT: 10.07 min Scan# 1358
Delta R.T. -0.02 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21




Instrument :
BNA_F
ClientSampled :


Tot Ion:165 Resp: 4691
Ion Ratio Lower Upper
165 100
63 12.8 36.4 54.6#
89 9.7 57.8 86.6#
182 11.1 1.6 2.4#



#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.42 min Scan# 1586
Delta R.T. 0.02 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21

Tgt Ion:188 Resp: 281168
Ion Ratio Lower Upper
188 100
94 21.0 8.5 12.7#
80 21.8 9.2 13.8#









#64
4,6-Dinitro-2-methylphenol
Concen: 9.545 ng
RT: 10.53 min Scan# 1436
Delta R.T. 0.03 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21




Instrument :
BNA_F
ClientSampleId :


Tot Ion:198 Resp: 953
Ion Ratio Lower Upper
198 100
51 845.0 37.7 77.7#
105 3129.4 36.3 76.3#



#68
Atrazine
Concen: 3.249 ng
RT: 11.11 min Scan# 1534
Delta R.T. 0.01 min
Lab File: BF102717.D
Acq: 3 Feb 2018 15:21

Tgt Ion:200 Resp: 9506
Ion Ratio Lower Upper
200 100
173 134.9 8.6 48.6#
215 86.1 25.1 65.1#





#74

Fluoranthene

Concen: 2.501 ng

RT: 12.63 min Scan# 1793

Delta R.T. 0.02 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampleId :


Tot Ion:202 Resp: 39409

Ion Ratio Lower Upper

202 100

101 15.6 0.0 34.1




203 34.6 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.04 min Scan# 2032

Delta R.T. 0.00 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion:240 Resp: 369742

Ion Ratio Lower Upper

240 100

120 15.0 9.5 14.3#



236 27.0 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 56.463 ng

RT: 13.00 min Scan# 1855

Delta R.T. 0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :

BNA_F

ClientSampled :


Tot Ion:244 Resp: 881703

Ion Ratio Lower Upper

244 100

212 6.9 5.4 8.0




122 13.5 11.0 16.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#79

Butylbenzylphthalate

Concen: 2.049 ng

RT: 13.46 min Scan# 1934

Delta R.T. 0.01 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21


Tgt Ion:149 Resp: 14513

Ion Ratio Lower Upper

149 100

91 82.2 56.8 85.2



206 10.5 14.1 21.1#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 2.287 ng
 RT: 14.02 min Scan# 2028
 Delta R.T. 0.00 min
 Lab File: BF102717.D
 Acq: 3 Feb 2018 15:21

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:149 Resp: 40612
 Ion Ratio Lower Upper
 149 100
 167 30.9 21.3 31.9
 279 4.3 3.0 4.4




Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 167.00 (166.70 to 167.70): E
 Ion 279.00 (278.70 to 279.70): E






#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.51 min Scan# 2282
 Delta R.T. 0.01 min
 Lab File: BF102717.D
 Acq: 3 Feb 2018 15:21

Tgt Ion:264 Resp: 395444
 Ion Ratio Lower Upper
 264 100
 260 24.9 19.2 28.8
 265 22.5 17.5 26.3



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 2.826 ng

RT: 15.07 min Scan# 2207

Delta R.T. -0.01 min





Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Instrument :
BNA_F
ClientSampleId :

Tot Ion:252 Resp: 72464

Ion	Ratio	Lower	Upper
252	100		
253	23.7	17.0	25.6
125	21.6	9.0	13.6#



#88

Benzo(k)fluoranthene

Concen: 2.958 ng

RT: 15.07 min Scan# 2207

Delta R.T. -0.04 min

Lab File: BF102717.D

Acq: 3 Feb 2018 15:21

Tgt Ion:252 Resp: 72464

Ion	Ratio	Lower	Upper
252	100		
253	23.7	17.9	26.9
125	21.6	10.2	15.2#

