


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_F\DATA\BF020320\  
 Data File : BF118790.D  
 Acq On : 3 Feb 2020 10:55  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF020320.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Feb 03 16:26:03 2020



AutoFind: Scans 1616, 1617, 1618; Background Corrected with Scan 1611

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 30.8      | 60744   | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 1324    | PASS             |
| 69          | 198          | 0.00         | 100          | 34.1      | 67146   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.7       | 502     | PASS             |
| 127         | 198          | 10           | 80           | 44.9      | 88482   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.2       | 361     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 197170  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 13467   | PASS             |
| 275         | 198          | 10           | 60           | 29.9      | 58997   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 6795    | PASS             |
| 441         | 198          | 0.01         | 100          | 12.8      | 25253   | PASS             |
| 442         | 198          | 50           | 100          | 80.8      | 159357  | PASS             |
| 443         | 442          | 15           | 24           | 19.3      | 30732   | PASS             |