

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020322\
 Data File : BF126788.D
 Acq On : 03 Feb 2022 09:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 11:31:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 15:08:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	40.303	-0.8	113	0.00
3	Pyridine	40.000	40.494	-1.2	112	0.00
4	n-Nitrosodimethylamine	40.000	40.164	-0.4	111	0.00
5 S	2-Fluorophenol	80.000	77.109	3.6	112	0.00
6	Aniline	40.000	40.383	-1.0	111	0.00
7 S	Phenol-d6	80.000	78.106	2.4	112	0.00
8	2-Chlorophenol	40.000	39.463	1.3	112	0.00
9	Benzaldehyde	40.000	38.561	3.6	112	0.00
10 C	Phenol	40.000	39.129	2.2	113	0.00
11	bis(2-Chloroethyl)ether	40.000	39.699	0.8	113	0.00
12	1,3-Dichlorobenzene	40.000	38.668	3.3	111	0.00
13 C	1,4-Dichlorobenzene	40.000	38.751	3.1	110	0.00
14	1,2-Dichlorobenzene	40.000	36.970	7.6	110	0.00
15	Benzyl Alcohol	40.000	39.592	1.0	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.849	0.4	111	0.00
17	2-Methylphenol	40.000	39.160	2.1	110	0.00
18	Hexachloroethane	40.000	38.793	3.0	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.220	4.5	109	0.00
20	3+4-Methylphenols	40.000	39.516	1.2	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	37.651	5.9	109	0.00
23 S	Nitrobenzene-d5	80.000	77.222	3.5	108	0.00
24	Nitrobenzene	40.000	38.859	2.9	109	0.00
25	Isophorone	40.000	38.785	3.0	109	0.00
26 C	2-Nitrophenol	40.000	40.159	-0.4	108	0.00
27	2,4-Dimethylphenol	40.000	39.403	1.5	110	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.998	2.5	109	0.00
29 C	2,4-Dichlorophenol	40.000	39.624	0.9	110	0.00
30	1,2,4-Trichlorobenzene	40.000	38.527	3.7	108	0.00
31	Naphthalene	40.000	38.616	3.5	110	0.00
32	Benzoic acid	40.000	33.864	15.3	89	0.00
33	4-Chloroaniline	40.000	39.270	1.8	110	0.00
34 C	Hexachlorobutadiene	40.000	38.145	4.6	109	0.00
35	Caprolactam	40.000	40.644	-1.6	111	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.172	2.1	111	0.00
37	2-Methylnaphthalene	40.000	39.015	2.5	110	0.00
38	1-Methylnaphthalene	40.000	39.031	2.4	110	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.392	4.0	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.253	4.4	104	0.00
42 S	2,4,6-Tribromophenol	80.000	77.961	2.5	108	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.862	2.8	110	0.00
44	2,4,5-Trichlorophenol	40.000	40.110	-0.3	112	0.00
45 S	2-Fluorobiphenyl	80.000	77.380	3.3	107	0.00
46	1,1'-Biphenyl	40.000	38.652	3.4	110	0.00
47	2-Chloronaphthalene	40.000	39.019	2.5	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020322\
 Data File : BF126788.D
 Acq On : 03 Feb 2022 09:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 11:31:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 15:08:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.935	0.2	110	0.00
49	Acenaphthylene	40.000	38.948	2.6	109	0.00
50	Dimethylphthalate	40.000	38.902	2.7	110	0.00
51	2,6-Dinitrotoluene	40.000	40.200	-0.5	111	0.00
52 C	Acenaphthene	40.000	38.734	3.2	108	0.00
53	3-Nitroaniline	40.000	40.862	-2.2	111	0.00
54 P	2,4-Dinitrophenol	40.000	41.136	-2.8	106	0.00
55	Dibenzofuran	40.000	38.023	4.9	108	0.00
56 P	4-Nitrophenol	40.000	40.572	-1.4	111	0.00
57	2,4-Dinitrotoluene	40.000	39.888	0.3	110	0.00
58	Fluorene	40.000	38.245	4.4	110	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.499	1.3	111	0.00
60	Diethylphthalate	40.000	38.247	4.4	109	0.00
61	4-Chlorophenyl-phenylether	40.000	38.174	4.6	108	0.00
62	4-Nitroaniline	40.000	40.752	-1.9	109	0.00
63	Azobenzene	40.000	38.870	2.8	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.739	-1.8	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.771	3.1	111	0.00
67	4-Bromophenyl-phenylether	40.000	38.849	2.9	108	0.00
68	Hexachlorobenzene	40.000	38.079	4.8	109	0.00
69	Atrazine	40.000	39.211	2.0	110	0.00
70 C	Pentachlorophenol	40.000	40.824	-2.1	110	0.00
71	Phenanthrene	40.000	37.932	5.2	109	0.00
72	Anthracene	40.000	37.964	5.1	110	0.00
73	Carbazole	40.000	37.454	6.4	110	0.00
74	Di-n-butylphthalate	40.000	37.988	5.0	109	0.00
75 C	Fluoranthene	40.000	37.979	5.1	111	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.00
77	Benzydine	40.000	41.423	-3.6	107	0.00
78	Pyrene	40.000	39.097	2.3	111	0.00
79 S	Terphenyl-d14	80.000	75.261	5.9	109	0.00
80	Butylbenzylphthalate	40.000	40.420	-1.1	113	0.00
81	Benzo(a)anthracene	40.000	39.414	1.5	115	0.00
82	3,3'-Dichlorobenzidine	40.000	41.300	-3.2	118	0.00
83	Chrysene	40.000	39.596	1.0	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.646	-1.6	115	0.00
85 c	Di-n-octyl phthalate	40.000	40.772	-1.9	120	0.00
86 I	Perylene-d12	20.000	20.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.528	-1.3	112	0.00
88	Benzo(b)fluoranthene	40.000	39.347	1.6	120	0.00
89	Benzo(k)fluoranthene	40.000	40.016	-0.0	119	0.00
90 C	Benzo(a)pyrene	40.000	39.490	1.3	118	0.00
91	Dibenzo(a,h)anthracene	40.000	40.263	-0.7	111	0.00
92	Benzo(g,h,i)perylene	40.000	41.084	-2.7	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020322\
Data File : BF126788.D
Acq On : 03 Feb 2022 09:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 03 11:31:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 02 15:08:47 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0