

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020322\
 Data File : BF126799.D
 Acq On : 03 Feb 2022 15:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 17:18:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 15:08:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	117	0.00
2	1,4-Dioxane	40.000	39.979	0.1	122	0.00
3	Pyridine	40.000	40.973	-2.4	123	0.00
4	n-Nitrosodimethylamine	40.000	40.220	-0.5	120	0.00
5 S	2-Fluorophenol	80.000	76.382	4.5	121	0.00
6	Aniline	40.000	40.586	-1.5	122	0.00
7 S	Phenol-d6	80.000	78.460	1.9	122	0.00
8	2-Chlorophenol	40.000	39.040	2.4	120	0.00
9	Benzaldehyde	40.000	38.759	3.1	122	0.00
10 C	Phenol	40.000	39.329	1.7	123	0.00
11	bis(2-Chloroethyl)ether	40.000	39.421	1.4	122	0.00
12	1,3-Dichlorobenzene	40.000	38.130	4.7	119	0.00
13 C	1,4-Dichlorobenzene	40.000	38.600	3.5	119	0.00
14	1,2-Dichlorobenzene	40.000	36.726	8.2	119	0.00
15	Benzyl Alcohol	40.000	39.159	2.1	117	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.017	-0.0	121	0.00
17	2-Methylphenol	40.000	39.410	1.5	121	0.00
18	Hexachloroethane	40.000	38.648	3.4	119	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.097	4.8	118	0.00
20	3+4-Methylphenols	40.000	39.058	2.4	120	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.00
22	Acetophenone	40.000	38.075	4.8	119	0.00
23 S	Nitrobenzene-d5	80.000	78.108	2.4	119	0.00
24	Nitrobenzene	40.000	38.997	2.5	119	0.00
25	Isophorone	40.000	39.535	1.2	121	0.00
26 C	2-Nitrophenol	40.000	41.668	-4.2	122	0.00
27	2,4-Dimethylphenol	40.000	40.233	-0.6	122	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.529	1.2	119	0.00
29 C	2,4-Dichlorophenol	40.000	40.005	-0.0	120	0.00
30	1,2,4-Trichlorobenzene	40.000	38.993	2.5	119	0.00
31	Naphthalene	40.000	38.683	3.3	119	0.00
32	Benzoic acid	40.000	43.120	-7.8	122	0.00
33	4-Chloroaniline	40.000	40.180	-0.4	122	0.00
34 C	Hexachlorobutadiene	40.000	38.417	4.0	119	0.00
35	Caprolactam	40.000	42.107	-5.3	124	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.460	1.3	121	0.00
37	2-Methylnaphthalene	40.000	39.068	2.3	120	0.00
38	1-Methylnaphthalene	40.000	39.164	2.1	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.587	6.0	117	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.922	2.7	116	0.00
42 S	2,4,6-Tribromophenol	80.000	78.833	1.5	121	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.474	3.8	120	0.00
44	2,4,5-Trichlorophenol	40.000	39.476	1.3	121	0.00
45 S	2-Fluorobiphenyl	80.000	76.346	4.6	117	0.00
46	1,1'-Biphenyl	40.000	38.574	3.6	121	0.00
47	2-Chloronaphthalene	40.000	38.687	3.3	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020322\
 Data File : BF126799.D
 Acq On : 03 Feb 2022 15:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 17:18:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 15:08:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.392	-1.0	122	0.00
49	Acenaphthylene	40.000	38.904	2.7	120	0.00
50	Dimethylphthalate	40.000	39.040	2.4	121	0.00
51	2,6-Dinitrotoluene	40.000	39.896	0.3	121	0.00
52 C	Acenaphthene	40.000	38.963	2.6	120	0.00
53	3-Nitroaniline	40.000	41.172	-2.9	123	0.00
54 P	2,4-Dinitrophenol	40.000	44.540	-11.3	127	0.00
55	Dibenzofuran	40.000	38.020	4.9	120	0.00
56 P	4-Nitrophenol	40.000	42.517	-6.3	128	0.00
57	2,4-Dinitrotoluene	40.000	40.648	-1.6	124	0.00
58	Fluorene	40.000	37.668	5.8	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.889	2.8	121	0.00
60	Diethylphthalate	40.000	38.137	4.7	119	0.00
61	4-Chlorophenyl-phenylether	40.000	37.814	5.5	118	0.00
62	4-Nitroaniline	40.000	42.584	-6.5	125	0.00
63	Azobenzene	40.000	39.015	2.5	121	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	117	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.101	-10.3	123	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.236	1.9	122	0.00
67	4-Bromophenyl-phenylether	40.000	39.624	0.9	120	0.00
68	Hexachlorobenzene	40.000	38.417	4.0	120	0.00
69	Atrazine	40.000	40.054	-0.1	121	0.00
70 C	Pentachlorophenol	40.000	42.415	-6.0	124	0.00
71	Phenanthrene	40.000	38.618	3.5	121	0.00
72	Anthracene	40.000	38.419	4.0	121	0.00
73	Carbazole	40.000	38.671	3.3	123	0.00
74	Di-n-butylphthalate	40.000	38.784	3.0	120	0.00
75 C	Fluoranthene	40.000	38.812	3.0	123	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	140	0.00
77	Benzydine	40.000	42.020	-5.1	134	0.00
78	Pyrene	40.000	35.435	11.4	125	0.00
79 S	Terphenyl-d14	80.000	67.739	15.3	122	0.00
80	Butylbenzylphthalate	40.000	38.025	4.9	132	0.00
81	Benzo(a)anthracene	40.000	38.987	2.5	141	0.00
82	3,3'-Dichlorobenzidine	40.000	43.509	-8.8	154	0.00
83	Chrysene	40.000	38.572	3.6	141	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.089	-0.2	141	0.00
85 c	Di-n-octyl phthalate	40.000	45.294	-13.2	165	0.00
86 I	Perylene-d12	20.000	20.000	0.0	153	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.416	11.5	131	0.00
88	Benzo(b)fluoranthene	40.000	39.197	2.0	160	0.00
89	Benzo(k)fluoranthene	40.000	40.563	-1.4	160	0.00
90 C	Benzo(a)pyrene	40.000	39.251	1.9	157	0.00
91	Dibenzo(a,h)anthracene	40.000	35.825	10.4	132	0.00
92	Benzo(g,h,i)perylene	40.000	35.209	12.0	129	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020322\
Data File : BF126799.D
Acq On : 03 Feb 2022 15:35
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 03 17:18:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 02 15:08:47 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0