

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020516\
 Data File : BF084878.D
 Acq On : 5 Feb 2016 23:47
 Operator : SJ/IZ
 Sample : H1311-02
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SUP-B007(10-11)

Quant Time: Feb 06 03:45:06 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 14:16:01 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	172487	20.00	ng	-0.02
21) Naphthalene-d8	7.95	136	705229	20.00	ng	-0.02
38) Acenaphthene-d10	9.70	164	318952	20.00	ng	-0.03
63) Phenanthrene-d10	11.17	188	526911	20.00	ng	-0.03
75) Chrysene-d12	13.80	240	372495	20.00	ng	-0.03
86) Perylene-d12	15.17	264	311085	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	1139412	102.89	ng	-0.02
7) Phenol-d6	6.32	99	1429824	104.15	ng	-0.02
23) Nitrobenzene-d5	7.23	82	850142	67.91	ng	-0.03
41) 2,4,6-Tribromophenol	10.49	330	371967	111.80	ng	-0.03
44) 2-Fluorobiphenyl	9.02	172	1396791	69.38	ng	-0.03
78) Terphenyl-d14	12.76	244	1137730	69.21	ng	-0.03
Target Compounds						Qvalue
22) Acetophenone	7.06	105	47202	2.54	ng	# 54
31) Naphthalene	7.97	128	188259	5.32	ng	98
37) 2-Methylnaphthalene	8.66	142	82305	3.72	ng	99
49) Dimethylphthalate	9.42	163	330202	13.59	ng	# 90
70) Phenanthrene	11.20	178	137516	4.71	ng	98
74) Fluoranthene	12.38	202	129573	4.38	ng	94
77) Pyrene	12.60	202	114441	4.01	ng	99
80) Benzo(a)anthracene	13.79	228	46625	2.02	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020516\
Data File : BF084878.D
Acq On : 5 Feb 2016 23:47
Operator : SJ/IZ
Sample : H1311-02
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SUP-B007(10-11)

Quant Time: Feb 06 03:45:06 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 03 14:16:01 2016
Response via : Initial Calibration

