

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020516\
 Data File : BF084883.D
 Acq On : 6 Feb 2016 2:05
 Operator : SJ/IZ
 Sample : H1311-11 2X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SUP-B006(0-2)

Manual Integrations
 APPROVED

UMANGI
 2/8/2016 3:03:15 PM

Quant Time: Feb 06 04:07:45 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 14:16:01 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	182069	20.00	ng	-0.02
21) Naphthalene-d8	7.95	136	737587	20.00	ng	-0.02
38) Acenaphthene-d10	9.70	164	326735	20.00	ng	-0.03
63) Phenanthrene-d10	11.17	188	577205	20.00	ng	-0.03
75) Chrysene-d12	13.80	240	402596	20.00	ng	-0.03
86) Perylene-d12	15.17	264	335437	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	578500	49.49	ng	-0.03
7) Phenol-d6	6.30	99	684573	47.24	ng	-0.03
23) Nitrobenzene-d5	7.23	82	520482	39.75	ng	-0.03
41) 2,4,6-Tribromophenol	10.49	330	167675	49.20	ng	-0.03
44) 2-Fluorobiphenyl	9.02	172	658614	27.06	ng	-0.03
78) Terphenyl-d14	12.76	244	470439	26.48	ng	-0.03
Target Compounds						Qvalue
49) Dimethylphthalate	9.42	163	133171	5.35	ng	# 90
70) Phenanthrene	11.20	178	195533	6.11	ng	99
74) Fluoranthene	12.37	202	182651	5.64	ng	99
77) Pyrene	12.60	202	182383	5.92	ng	99
80) Benzo(a)anthracene	13.79	228	78541	3.14	ng	98
82) Chrysene	13.83	228	50932	2.10	ng	97
87) Benzo(b)fluoranthene	14.80	252	60664m	2.69	ng	
89) Benzo(a)pyrene	15.12	252	46857	2.44	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020516\
Data File : BF084883.D
Acq On : 6 Feb 2016 2:05
Operator : SJ/IZ
Sample : H1311-11 2X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SUP-B006(0-2)

Manual Integrations
APPROVED

UMANGI
2/8/2016 3:03:15 PM

Quant Time: Feb 06 04:07:45 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 03 14:16:01 2016
Response via : Initial Calibration

