

Data Path : U:\HPCHEM1\BNA F\DATA\BF020518\
 Data File : BF102735.D
 Acq On : 5 Feb 2018 14:01
 Operator : SJ/JU
 Sample : J1424-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 IRRC-2-COMP

Quant Time: Feb 05 16:37:31 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 12:44:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	225440	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	920143	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	401726	20.00	ng	0.00
63) Phenanthrene-d10	11.40	188	630130	20.00	ng	0.00
75) Chrysene-d12	14.04	240	405899	20.00	ng	-0.01
86) Perylene-d12	15.52	264	377003	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1519719	102.38	ng	0.01
7) Phenol-d6	6.51	99	1831952	102.49	ng	0.00
23) Nitrobenzene-d5	7.45	82	1132185	85.36	ng	0.00
41) 2,4,6-Tribromophenol	10.71	330	344012	106.39	ng	0.00
44) 2-Fluorobiphenyl	9.24	172	1679446	69.37	ng	0.00
78) Terphenyl-d14	12.99	244	1213088	70.76	ng	0.00
Target Compounds						
10) Phenol	6.52	94	68920	3.598	ng	89
49) Dimethylphthalate	9.63	163	107958	3.447	ng	99
70) Phenanthrene	11.43	178	78541	2.238	ng	97
74) Fluoranthene	12.62	202	137740	3.901	ng	99
77) Pyrene	12.84	202	138550	4.353	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF020518\
Data File : BF102735.D
Acq On : 5 Feb 2018 14:01
Operator : SJ/JU
Sample : J1424-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
IRRC-2-COMP

Quant Time: Feb 05 16:37:31 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 05 12:44:16 2018
Response via : Initial Calibration

