

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/5/2024	BNA_F	<u>BF137268.D</u>
Compound Name	Response	Retention Time
DDT	220927	13.492
DDD	3168	13.221
DDE	193	12.857
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3361	224288	1.50