

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/6/2024	BNA_F	<u>BF137289.D</u>
Compound Name	Response	Retention Time
DDT	228705	13.498
DDD	3229	13.222
DDE	54	12.857
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3283	231988	1.42