

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020725\
 Data File : BF141506.D
 Acq On : 07 Feb 2025 14:33
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 02/10/2025
 Supervised By :mohammad ahmed 02/12/2025

Quant Time: Feb 07 15:09:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 16:58:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.787	152	91880	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	360290	20.000	ng	0.00
39) Acenaphthene-d10	9.828	164	198348	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	340922	20.000	ng	0.00
76) Chrysene-d12	13.957	240	238177	20.000	ng	0.00
86) Perylene-d12	15.416	264	192868	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.404	112	475650	81.160	ng	0.00
7) Phenol-d6	6.428	99	587648	79.333	ng	-0.01
23) Nitrobenzene-d5	7.357	82	553575	80.140	ng	0.00
42) 2,4,6-Tribromophenol	10.616	330	158428	79.512	ng	-0.01
45) 2-Fluorobiphenyl	9.151	172	1006130	78.150	ng	0.00
79) Terphenyl-d14	12.904	244	1115806	79.087	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.475	88	94001	39.852	ng	100
3) Pyridine	3.199	79	234861	41.842	ng	99
4) n-Nitrosodimethylamine	3.157	42	146825	39.505	ng	96
6) Aniline	6.451	93	276601	39.807	ng	94
8) 2-Chlorophenol	6.575	128	254071	40.121	ng	96
9) Benzaldehyde	6.340	77	155074	39.396	ng	100
10) Phenol	6.446	94	307937	39.942	ng	99
11) bis(2-Chloroethyl)ether	6.528	93	227223	39.239	ng	99
12) 1,3-Dichlorobenzene	6.728	146	265506	39.713	ng	98
13) 1,4-Dichlorobenzene	6.804	146	265354	38.847	ng	98
14) 1,2-Dichlorobenzene	6.963	146	252127	39.770	ng	100
15) Benzyl Alcohol	6.934	79	221923	39.069	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.069	45	389773	39.320	ng	99
17) 2-Methylphenol	7.051	107	195478	39.445	ng	99
18) Hexachloroethane	7.298	117	100140	39.613	ng	99
19) n-Nitroso-di-n-propyla...	7.204	70	179697	40.556	ng	97
20) 3+4-Methylphenols	7.204	107	244214	38.777	ng	98
22) Acetophenone	7.198	105	352167	39.294	ng	99
24) Nitrobenzene	7.375	77	268254	39.825	ng	100
25) Isophorone	7.610	82	437623	39.733	ng	100
26) 2-Nitrophenol	7.687	139	137445	41.014	ng	99
27) 2,4-Dimethylphenol	7.728	122	158012	40.362	ng	99
28) bis(2-Chloroethoxy)met...	7.822	93	285112	40.398	ng	100
29) 2,4-Dichlorophenol	7.940	162	213366	40.337	ng	99
30) 1,2,4-Trichlorobenzene	8.016	180	229655	39.869	ng	99
31) Naphthalene	8.092	128	749587	39.968	ng	100
32) Benzoic acid	7.857	122	163033	40.092	ng	98
33) 4-Chloroaniline	8.145	127	235464	35.960	ng	100
34) Hexachlorobutadiene	8.210	225	137788	39.845	ng	99
35) Caprolactam	8.516	113	67233	41.746	ng	97
36) 4-Chloro-3-methylphenol	8.634	107	234989	40.105	ng	100
37) 2-Methylnaphthalene	8.787	142	481204	39.430	ng	100
38) 1-Methylnaphthalene	8.881	142	465894	39.416	ng	97
40) 1,2,4,5-Tetrachloroben...	8.951	216	226553	39.480	ng	99
41) Hexachlorocyclopentadiene	8.934	237	76451	40.055	ng	99
43) 2,4,6-Trichlorophenol	9.063	196	145794	39.310	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020725\
 Data File : BF141506.D
 Acq On : 07 Feb 2025 14:33
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Feb 07 15:09:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 16:58:59 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 02/10/2025
 Supervised By :mohammad ahmed 02/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.110	196	158176	39.352	ng	99
46) 1,1'-Biphenyl	9.251	154	606143	39.417	ng	100
47) 2-Chloronaphthalene	9.275	162	448999	39.485	ng	100
48) 2-Nitroaniline	9.375	65	152856	40.054	ng	98
49) Acenaphthylene	9.686	152	673155	39.800	ng	99
50) Dimethylphthalate	9.551	163	530887	39.426	ng	99
51) 2,6-Dinitrotoluene	9.616	165	116863	39.701	ng	99
52) Acenaphthene	9.863	154	442729m	38.936	ng	
53) 3-Nitroaniline	9.786	138	124846	39.406	ng	99
54) 2,4-Dinitrophenol	9.892	184	67053	38.328	ng	97
55) Dibenzofuran	10.034	168	656713	39.347	ng	100
56) 4-Nitrophenol	9.957	139	95307	40.524	ng	97
57) 2,4-Dinitrotoluene	10.016	165	158641	40.483	ng	97
58) Fluorene	10.375	166	507132	39.298	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.151	232	137272	39.668	ng	97
60) Diethylphthalate	10.251	149	520921	39.320	ng	100
61) 4-Chlorophenyl-phenyle...	10.369	204	244431	39.371	ng	99
62) 4-Nitroaniline	10.398	138	129336	40.203	ng	97
63) Azobenzene	10.528	77	516030	39.181	ng	99
65) 4,6-Dinitro-2-methylph...	10.428	198	94709	39.992	ng	98
66) n-Nitrosodiphenylamine	10.486	169	430480	39.446	ng	99
67) 4-Bromophenyl-phenylether	10.857	248	148075	38.862	ng	96
68) Hexachlorobenzene	10.922	284	156412	39.865	ng	98
69) Atrazine	11.016	200	124255	37.952	ng	99
70) Pentachlorophenol	11.122	266	100181	39.932	ng	99
71) Phenanthrene	11.339	178	737878	39.319	ng	99
72) Anthracene	11.392	178	746361	39.564	ng	100
73) Carbazole	11.545	167	684333	40.316	ng	100
74) Di-n-butylphthalate	11.880	149	785155	40.060	ng	100
75) Fluoranthene	12.527	202	796201	40.018	ng	99
77) Benzidine	12.651	184	232310	44.226	ng	99
78) Pyrene	12.757	202	810550	39.977	ng	100
80) Butylbenzylphthalate	13.374	149	305087	43.443	ng	99
81) Benzo(a)anthracene	13.945	228	626940	39.598	ng	99
82) 3,3'-Dichlorobenzidine	13.910	252	183288	40.414	ng	98
83) Chrysene	13.986	228	556110	38.041	ng	99
84) Bis(2-ethylhexyl)phtha...	13.939	149	361384	45.626	ng	99
85) Di-n-octyl phthalate	14.563	149	484418	42.871	ng	99
87) Indeno(1,2,3-cd)pyrene	16.862	276	480414	40.313	ng	100
88) Benzo(b)fluoranthene	14.998	252	549964	42.468	ng	99
89) Benzo(k)fluoranthene	15.027	252	412835	37.333	ng	99
90) Benzo(a)pyrene	15.357	252	420423	40.121	ng	99
91) Dibenzo(a,h)anthracene	16.880	278	390089	40.397	ng	100
92) Benzo(g,h,i)perylene	17.298	276	396572	39.245	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Reviewed By :Anahy Claudio 02/10/2025
Supervised By :mohammad ahmed 02/12/2025

[illegible]