


Data Path : Z:\HPCHEM1\BNA_F\Data\BF020817\
 Data File : BF092853.D
 Acq On : 8 Feb 2017 14:47
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 30 14:50:05 2017



AutoFind: Scans 838, 839, 840; Background Corrected with Scan 835

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.9	53983	PASS
68	69	0.00	2	2.0	891	PASS
69	198	0.00	100	40.8	45036	PASS
70	69	0.00	2	0.4	201	PASS
127	198	10	80	48.0	52959	PASS
197	198	0.00	2	1.0	1063	PASS
198	198	100	100	100.0	110421	PASS
199	198	5	9	6.5	7129	PASS
275	198	10	60	22.1	24436	PASS
365	198	1	100	2.2	2440	PASS
441	442	0.01	24	16.1	13676	PASS
442	198	50	100	76.7	84686	PASS
443	442	15	24	18.4	15574	PASS