


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020817\
 Data File : BF092878.D
 Acq On : 9 Feb 2017 10:21
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 30 14:50:05 2017



AutoFind: Scans 838, 839, 840; Background Corrected with Scan 835

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.0	34861	PASS
68	69	0.00	2	2.0	582	PASS
69	198	0.00	100	36.6	29699	PASS
70	69	0.00	2	0.4	119	PASS
127	198	10	80	45.9	37269	PASS
197	198	0.00	2	1.1	914	PASS
198	198	100	100	100.0	81124	PASS
199	198	5	9	6.6	5352	PASS
275	198	10	60	22.1	17952	PASS
365	198	1	100	2.0	1601	PASS
441	442	0.01	24	15.7	9815	PASS
442	198	50	100	76.8	62327	PASS
443	442	15	24	18.8	11688	PASS