

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020818\
 Data File : BF102859.D
 Acq On : 8 Feb 2018 16:39
 Operator : SJ/JU
 Sample : J1470-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 3N-6

Manual Integrations
 APPROVED

Sohil
 2/9/2018 4:33:31 PM

Quant Time: Feb 08 18:27:30 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 12:44:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	187530	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	730293	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	291357	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	393093	20.00	ng	0.00
75) Chrysene-d12	14.05	240	339093	20.00	ng	0.00
86) Perylene-d12	15.54	264	298876	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1083885	87.78	ng	0.01
7) Phenol-d6	6.51	99	1291076	86.83	ng	0.00
23) Nitrobenzene-d5	7.45	82	809234	76.87	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	161741	68.97	ng	0.00
44) 2-Fluorobiphenyl	9.24	172	1023225	58.28	ng	0.00
78) Terphenyl-d14	13.00	244	589407	41.16	ng	0.00
Target Compounds						
10) Phenol	6.52	94	32598	2.046	ng	89
20) 3+4-Methylphenols	7.28	107	40494	2.800	ng	88
49) Dimethylphthalate	9.63	163	102802	4.525	ng	99
59) Diethylphthalate	10.33	149	241969	10.787	ng	99
79) Butylbenzylphthalate	13.47	149	47775m	4.701	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF020818\
Data File : BF102859.D
Acq On : 8 Feb 2018 16:39
Operator : SJ/JU
Sample : J1470-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
3N-6

Manual Integrations
APPROVED

Sohil
2/9/2018 4:33:31 PM

Quant Time: Feb 08 18:27:30 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 05 12:44:16 2018
Response via : Initial Calibration

