

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137341.D
Acq On : 08 Feb 2024 16:52
Operator : CG\JU
Sample : P1342-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MR-323P-GW-20240131-0

Quant Time: Feb 08 17:19:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	114712	20.000	ng	-0.01
21) Naphthalene-d8	8.063	136	433804	20.000	ng	#-0.01
39) Acenaphthene-d10	9.816	164	232518	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	379654	20.000	ng	0.00
76) Chrysene-d12	13.962	240	256035	20.000	ng	-0.01
86) Perylene-d12	15.445	264	302562	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	365513	52.577	ng	0.00
7) Phenol-d6	6.416	99	286102	28.838	ng	-0.01
23) Nitrobenzene-d5	7.345	82	767803	79.243	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	283688	108.033	ng	0.00
45) 2-Fluorobiphenyl	9.145	172	1278998	75.616	ng	0.00
79) Terphenyl-d14	12.910	244	1165639	61.891	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137341.D
Acq On : 08 Feb 2024 16:52
Operator : CG\JU
Sample : P1342-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MR-323P-GW-20240131-0

Quant Time: Feb 08 17:19:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration

