

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
 Data File : BF137347.D
 Acq On : 08 Feb 2024 19:48
 Operator : CG\JU
 Sample : P1342-09DL 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MR-317AP-GW-20240131-ODL

Quant Time: Feb 09 01:38:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 04:56:22 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 02/09/2024
 Supervised By :mohammad ahmed 02/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	83987	20.000	ng	-0.01
21) Naphthalene-d8	8.063	136	325687	20.000	ng	-0.01
39) Acenaphthene-d10	9.816	164	543563	20.000	ng	#-0.01
64) Phenanthrene-d10	11.310	188	274497	20.000	ng	0.00
76) Chrysene-d12	13.962	240	183447	20.000	ng	#-0.01
86) Perylene-d12	15.445	264	220963	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.410	112	33174	6.518	ng	0.00
7) Phenol-d6	6.422	99	22881	3.150	ng	0.00
23) Nitrobenzene-d5	7.339	82	71168	9.783	ng	-0.02
42) 2,4,6-Tribromophenol	10.610	330	29292	4.772	ng	-0.01
45) 2-Fluorobiphenyl	9.139	172	147098	3.720	ng	-0.01
79) Terphenyl-d14	12.904	244	122879	9.106	ng	-0.02
Target Compounds						
15) Benzyl Alcohol	6.934	79	444444	86.124	ng	Qvalue 94
27) 2,4-Dimethylphenol	7.675	122	192024m	50.034	ng	
32) Benzoic acid	7.904	122	454544	139.621	ng	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137347.D
Acq On : 08 Feb 2024 19:48
Operator : CG\JU
Sample : P1342-09DL 10X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MR-317AP-GW-20240131-ODL

Quant Time: Feb 09 01:38:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 02/09/2024
Supervised By :mohammad ahmed 02/09/2024

