

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137348.D
Acq On : 08 Feb 2024 20:17
Operator : CG\JU
Sample : P1367-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EB-6-0-2

Quant Time: Feb 09 01:39:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.787	152	105549	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	400704	20.000	ng	-0.01
39) Acenaphthene-d10	9.816	164	208654	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	353297	20.000	ng	0.00
76) Chrysene-d12	13.963	240	240768	20.000	ng	#-0.01
86) Perylene-d12	15.445	264	269269	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	878852	137.394	ng	0.02
7) Phenol-d6	6.428	99	1036565	113.552	ng	0.00
23) Nitrobenzene-d5	7.351	82	797794	89.139	ng	0.00
42) 2,4,6-Tribromophenol	10.610	330	287018	121.802	ng	-0.01
45) 2-Fluorobiphenyl	9.145	172	1286494	84.758	ng	0.00
79) Terphenyl-d14	12.910	244	1278289	72.176	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137348.D
Acq On : 08 Feb 2024 20:17
Operator : CG\JU
Sample : P1367-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EB-6-0-2

Quant Time: Feb 09 01:39:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration

