


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137333.D
Acq On : 08 Feb 2024 12:21
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Jan 31 04:56:22 2024






AutoFind: Scans 1612, 1613, 1614; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	14348	PASS
68	69	0.00	2	0.7	99	PASS
69	198	0.00	100	37.0	13959	PASS
70	69	0.00	2	0.4	56	PASS
127	198	10	80	42.9	16207	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	37736	PASS
199	198	5	9	6.8	2548	PASS
275	198	10	60	27.1	10212	PASS
365	198	1	100	3.4	1298	PASS
441	198	0.01	100	14.5	5484	PASS
442	442	50	100	100.0	35126	PASS
443	442	15	24	19.5	6837	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137333.D
Acq On : 08 Feb 2024 12:21
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 08 13:27:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration



TIC: BF137333.D\data.ms

(70) Pentachlorophenol (C)

11.110min (-0.011) 39858.71 ng




response 30308

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.30	58.73
264.00	61.00	63.33
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020824\
Data File : BF137333.D
Acq On : 08 Feb 2024 12:21
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 08 13:27:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration



TIC: BF137333.D\data.ms

(77) Benzidine

12.663min (-0.005) 0.00 ng

response 128586

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	15.43
183.00	11.00	11.92
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/8/2024	BNA_F	<u>BF137333.D</u>
Compound Name	Response	Retention Time
DDT	78174	13.486
DDD	1716	13.186
DDE	0	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1716	79890	2.15

Instrument :
BNA_F
ClientSampleId :
DFTPP