

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020916\  
 Data File : BF084970.D  
 Acq On : 9 Feb 2016 23:39  
 Operator : SJ/IZ  
 Sample : H1377-04  
 Misc :  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SB03-0-1

Quant Time: Feb 10 05:02:32 2016  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Feb 03 14:16:01 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.64  | 152  | 163406   | 20.00  | ng    | -0.05    |
| 21) Naphthalene-d8          | 7.92  | 136  | 694893   | 20.00  | ng    | -0.06    |
| 38) Acenaphthene-d10        | 9.66  | 164  | 326245   | 20.00  | ng    | -0.07    |
| 63) Phenanthrene-d10        | 11.14 | 188  | 581988   | 20.00  | ng    | -0.07    |
| 75) Chrysene-d12            | 13.77 | 240  | 408366   | 20.00  | ng    | -0.07    |
| 86) Perylene-d12            | 15.13 | 264  | 324379   | 20.00  | ng    | -0.08    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.22  | 112  | 1020332  | 97.26  | ng    | -0.06    |
| 7) Phenol-d6                | 6.29  | 99   | 1397636  | 107.46 | ng    | -0.05    |
| 23) Nitrobenzene-d5         | 7.21  | 82   | 917735   | 74.40  | ng    | -0.06    |
| 41) 2,4,6-Tribromophenol    | 10.45 | 330  | 314146   | 92.31  | ng    | -0.07    |
| 44) 2-Fluorobiphenyl        | 9.00  | 172  | 1435606  | 69.84  | ng    | -0.06    |
| 78) Terphenyl-d14           | 12.73 | 244  | 1095222  | 60.77  | ng    | -0.07    |
| Target Compounds            |       |      |          |        |       | Qvalue   |
| 49) Dimethylphthalate       | 9.40  | 163  | 256968   | 10.34  | ng    | # 90     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020916\  
Data File : BF084970.D  
Acq On : 9 Feb 2016 23:39  
Operator : SJ/IZ  
Sample : H1377-04  
Misc :  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SB03-0-1

Quant Time: Feb 10 05:02:32 2016  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Feb 03 14:16:01 2016  
Response via : Initial Calibration

