

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020922\
 Data File : BF126831.D
 Acq On : 09 Feb 2022 17:38
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF020922

Quant Time: Feb 10 00:48:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	0.000	100.0#	0	-2.79#
3	Pyridine	40.000	39.127	2.2	96	0.00
4	n-Nitrosodimethylamine	40.000	39.137	2.2	97	0.00
5 S	2-Fluorophenol	80.000	78.198	2.3	98	0.00
6	Aniline	40.000	39.116	2.2	96	0.00
7 S	Phenol-d6	80.000	78.571	1.8	97	0.00
8	2-Chlorophenol	40.000	39.331	1.7	96	0.00
9	Benzaldehyde	40.000	37.068	7.3	98	0.00
10 C	Phenol	40.000	39.016	2.5	97	0.00
11	bis(2-Chloroethyl)ether	40.000	39.127	2.2	97	0.00
12	1,3-Dichlorobenzene	40.000	38.730	3.2	96	0.00
13 C	1,4-Dichlorobenzene	40.000	39.058	2.4	97	0.00
14	1,2-Dichlorobenzene	40.000	38.761	3.1	96	0.00
15	Benzyl Alcohol	40.000	40.520	-1.3	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.750	3.1	95	0.00
17	2-Methylphenol	40.000	39.251	1.9	96	0.00
18	Hexachloroethane	40.000	39.231	1.9	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.690	0.8	99	0.00
20	3+4-Methylphenols	40.000	39.399	1.5	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	38.269	4.3	97	0.00
23 S	Nitrobenzene-d5	80.000	78.565	1.8	98	0.00
24	Nitrobenzene	40.000	39.857	0.4	99	0.00
25	Isophorone	40.000	38.376	4.1	95	0.00
26 C	2-Nitrophenol	40.000	40.783	-2.0	98	0.00
27	2,4-Dimethylphenol	40.000	39.662	0.8	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.221	4.4	95	0.00
29 C	2,4-Dichlorophenol	40.000	38.666	3.3	97	0.00
30	1,2,4-Trichlorobenzene	40.000	38.101	4.7	95	0.00
31	Naphthalene	40.000	38.199	4.5	96	0.00
32	Benzoic acid	40.000	42.022	-5.1	96	0.00
33	4-Chloroaniline	40.000	38.479	3.8	95	0.00
34 C	Hexachlorobutadiene	40.000	38.634	3.4	99	0.00
35	Caprolactam	40.000	39.820	0.4	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.071	2.3	96	0.00
37	2-Methylnaphthalene	40.000	38.286	4.3	96	0.00
38	1-Methylnaphthalene	40.000	38.291	4.3	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.122	4.7	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.208	-0.5	94	0.00
42 S	2,4,6-Tribromophenol	80.000	78.535	1.8	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.947	2.6	97	0.00
44	2,4,5-Trichlorophenol	40.000	39.509	1.2	97	0.00
45 S	2-Fluorobiphenyl	80.000	75.692	5.4	97	0.00
46	1,1'-Biphenyl	40.000	38.256	4.4	97	0.00
47	2-Chloronaphthalene	40.000	38.335	4.2	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020922\
 Data File : BF126831.D
 Acq On : 09 Feb 2022 17:38
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF020922

Quant Time: Feb 10 00:48:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.418	-1.0	96	0.00
49	Acenaphthylene	40.000	38.788	3.0	96	0.00
50	Dimethylphthalate	40.000	38.390	4.0	96	0.00
51	2,6-Dinitrotoluene	40.000	39.613	1.0	95	0.00
52 C	Acenaphthene	40.000	39.000	2.5	96	0.00
53	3-Nitroaniline	40.000	40.422	-1.1	98	0.00
54 P	2,4-Dinitrophenol	40.000	40.538	-1.3	99	0.00
55	Dibenzofuran	40.000	38.769	3.1	97	0.00
56 P	4-Nitrophenol	40.000	42.359	-5.9	98	0.00
57	2,4-Dinitrotoluene	40.000	40.496	-1.2	97	0.00
58	Fluorene	40.000	38.154	4.6	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.420	1.4	98	0.00
60	Diethylphthalate	40.000	38.855	2.9	96	0.00
61	4-Chlorophenyl-phenylether	40.000	38.448	3.9	98	0.00
62	4-Nitroaniline	40.000	40.947	-2.4	96	0.00
63	Azobenzene	40.000	38.289	4.3	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.029	-2.6	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.379	6.6	95	0.00
67	4-Bromophenyl-phenylether	40.000	37.906	5.2	96	0.00
68	Hexachlorobenzene	40.000	37.517	6.2	95	0.00
69	Atrazine	40.000	39.065	2.3	95	0.00
70 C	Pentachlorophenol	40.000	41.861	-4.7	98	0.00
71	Phenanthrene	40.000	37.732	5.7	96	0.00
72	Anthracene	40.000	37.967	5.1	96	0.00
73	Carbazole	40.000	38.258	4.4	97	0.00
74	Di-n-butylphthalate	40.000	38.419	4.0	96	0.00
75 C	Fluoranthene	40.000	38.511	3.7	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	38.140	4.6	90	0.00
78	Pyrene	40.000	38.551	3.6	97	0.00
79 S	Terphenyl-d14	80.000	76.137	4.8	98	0.00
80	Butylbenzylphthalate	40.000	40.485	-1.2	99	0.00
81	Benzo(a)anthracene	40.000	38.421	3.9	98	0.00
82	3,3'-Dichlorobenzidine	40.000	38.979	2.6	98	0.00
83	Chrysene	40.000	38.816	3.0	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.023	2.4	100	0.00
85 c	Di-n-octyl phthalate	40.000	38.906	2.7	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.438	-3.6	97	0.00
88	Benzo(b)fluoranthene	40.000	39.939	0.2	99	0.00
89	Benzo(k)fluoranthene	40.000	39.056	2.4	97	0.00
90 C	Benzo(a)pyrene	40.000	39.550	1.1	96	0.00
91	Dibenzo(a,h)anthracene	40.000	41.564	-3.9	98	0.00
92	Benzo(g,h,i)perylene	40.000	41.454	-3.6	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020922\
Data File : BF126831.D
Acq On : 09 Feb 2022 17:38
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF020922

Quant Time: Feb 10 00:48:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0