

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021017\
 Data File : BF092901.D
 Acq On : 10 Feb 2017 23:34
 Operator : SJ/MA
 Sample : I1777-08
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 38-SB-01-0-2

Manual Integrations
 APPROVED

mohammad
 2/13/2017 12:41:11 PM

Quant Time: Feb 11 00:36:12 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 10 23:31:14 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	163023	20.00	ng	-0.01
21) Naphthalene-d8	8.18	136	702071	20.00	ng	-0.01
38) Acenaphthene-d10	9.93	164	402935	20.00	ng	-0.01
63) Phenanthrene-d10	11.41	188	682077	20.00	ng	-0.01
75) Chrysene-d12	14.05	240	494566	20.00	ng	-0.01
86) Perylene-d12	15.51	264	381669	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1191540	125.40	ng	0.01
7) Phenol-d6	6.53	99	1302728	106.55	ng	0.00
23) Nitrobenzene-d5	7.46	82	869319	68.95	ng	-0.01
41) 2,4,6-Tribromophenol	10.73	330	334428	114.76	ng	-0.01
44) 2-Fluorobiphenyl	9.25	172	1511264	67.95	ng	-0.01
78) Terphenyl-d14	13.00	244	1383855	65.75	ng	-0.01
Target Compounds						
10) Phenol	6.54	94	94170	6.58	ng	Qvalue # 73
49) Dimethylphthalate	9.65	163	132058	4.87	ng	98
74) Fluoranthene	12.63	202	164061	4.07	ng	97
77) Pyrene	12.85	202	151317	4.09	ng	100
80) Benzo(a)anthracene	14.04	228	81902	2.71	ng	98
82) Chrysene	14.08	228	61650m	2.18	ng	
87) Benzo(b)fluoranthene	15.08	252	67029m	2.67	ng	
89) Benzo(a)pyrene	15.44	252	49854	2.29	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021017\
Data File : BF092901.D
Acq On : 10 Feb 2017 23:34
Operator : SJ/MA
Sample : I1777-08
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
38-SB-01-0-2

Manual Integrations
APPROVED

mohammad
2/13/2017 12:41:11 PM

Quant Time: Feb 11 00:36:12 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 10 23:31:14 2017
Response via : Initial Calibration

