




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021017\
Data File : BF092906.D
Acq On : 11 Feb 2017 4:42
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 11 05:45:35 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 30 14:50:05 2017
Response via : Initial Calibration



TIC: BF092906.D

(69) Pentachlorophenol (C)




11.231min (-0.068) 69813.26ng

response 115705

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	64.56
264.00	62.90	64.10
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021017\
 Data File : BF092906.D
 Acq On : 11 Feb 2017 4:42
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 11 05:45:35 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 30 14:50:05 2017
 Response via : Initial Calibration



TIC: BF092906.D

(76) Benzidine

12.762min (-0.069) 0.00ng m

response 672835

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	14.69
183.00	11.50	12.61
0.00	0.00	0.00