

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021022\
 Data File : BF126842.D
 Acq On : 10 Feb 2022 14:33
 Operator : CG\JU
 Sample : N1432-14
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-5

Quant Time: Feb 10 15:11:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.939	152	121185	20.000	ng	0.00
21) Naphthalene-d8	8.222	136	482720	20.000	ng	0.00
39) Acenaphthene-d10	9.980	164	249617	20.000	ng	0.00
64) Phenanthrene-d10	11.469	188	445729	20.000	ng	0.00
76) Chrysene-d12	14.110	240	268912	20.000	ng	-0.01
86) Perylene-d12	15.621	264	197938	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.569	112	960158	121.988	ng	0.01
7) Phenol-d6	6.563	99	1174936	110.576	ng	0.00
23) Nitrobenzene-d5	7.504	82	930179	88.910	ng	0.00
42) 2,4,6-Tribromophenol	10.775	330	354500	132.470	ng	0.00
45) 2-Fluorobiphenyl	9.298	172	1429421	86.200	ng	0.00
79) Terphenyl-d14	13.063	244	1714558	91.403	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.686	163	51793	2.932	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021022\
Data File : BF126842.D
Acq On : 10 Feb 2022 14:33
Operator : CG\JU
Sample : N1432-14
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-5

Quant Time: Feb 10 15:11:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

