

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021022\
 Data File : BF126834.D
 Acq On : 10 Feb 2022 10:35
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 10 12:45:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	0.000	100.0#	0	-2.79#
3	Pyridine	40.000	39.463	1.3	106	0.00
4	n-Nitrosodimethylamine	40.000	39.987	0.0	108	0.00
5 S	2-Fluorophenol	80.000	77.870	2.7	106	0.00
6	Aniline	40.000	39.618	1.0	106	0.00
7 S	Phenol-d6	80.000	78.372	2.0	106	0.00
8	2-Chlorophenol	40.000	39.525	1.2	106	0.00
9	Benzaldehyde	40.000	36.468	8.8	105	0.00
10 C	Phenol	40.000	39.116	2.2	106	0.00
11	bis(2-Chloroethyl)ether	40.000	39.020	2.4	105	0.00
12	1,3-Dichlorobenzene	40.000	38.925	2.7	105	0.00
13 C	1,4-Dichlorobenzene	40.000	38.796	3.0	105	0.00
14	1,2-Dichlorobenzene	40.000	38.813	3.0	105	0.00
15	Benzyl Alcohol	40.000	40.502	-1.3	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.902	2.7	104	0.00
17	2-Methylphenol	40.000	39.440	1.4	106	0.00
18	Hexachloroethane	40.000	39.251	1.9	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.882	2.8	106	0.00
20	3+4-Methylphenols	40.000	39.187	2.0	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	38.098	4.8	105	0.00
23 S	Nitrobenzene-d5	80.000	78.781	1.5	107	0.00
24	Nitrobenzene	40.000	39.304	1.7	106	0.00
25	Isophorone	40.000	38.453	3.9	104	0.00
26 C	2-Nitrophenol	40.000	40.625	-1.6	107	0.00
27	2,4-Dimethylphenol	40.000	39.164	2.1	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.169	4.6	104	0.00
29 C	2,4-Dichlorophenol	40.000	38.635	3.4	105	0.00
30	1,2,4-Trichlorobenzene	40.000	38.612	3.5	105	0.00
31	Naphthalene	40.000	38.130	4.7	105	0.00
32	Benzoic acid	40.000	42.562	-6.4	106	0.00
33	4-Chloroaniline	40.000	39.093	2.3	106	0.00
34 C	Hexachlorobutadiene	40.000	38.594	3.5	108	0.00
35	Caprolactam	40.000	40.344	-0.9	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.886	2.8	104	0.00
37	2-Methylnaphthalene	40.000	38.267	4.3	105	0.00
38	1-Methylnaphthalene	40.000	37.853	5.4	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.789	5.5	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.038	2.4	99	0.00
42 S	2,4,6-Tribromophenol	80.000	76.138	4.8	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.487	1.3	106	0.00
44	2,4,5-Trichlorophenol	40.000	39.191	2.0	105	0.00
45 S	2-Fluorobiphenyl	80.000	75.417	5.7	104	0.00
46	1,1'-Biphenyl	40.000	37.964	5.1	104	0.00
47	2-Chloronaphthalene	40.000	38.338	4.2	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021022\
 Data File : BF126834.D
 Acq On : 10 Feb 2022 10:35
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 10 12:45:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.325	-0.8	104	0.00
49	Acenaphthylene	40.000	38.571	3.6	103	0.00
50	Dimethylphthalate	40.000	37.421	6.4	102	0.00
51	2,6-Dinitrotoluene	40.000	39.782	0.5	104	0.00
52 C	Acenaphthene	40.000	38.493	3.8	103	0.00
53	3-Nitroaniline	40.000	39.698	0.8	105	0.00
54 P	2,4-Dinitrophenol	40.000	37.724	5.7	99	0.00
55	Dibenzofuran	40.000	38.003	5.0	103	0.00
56 P	4-Nitrophenol	40.000	41.175	-2.9	103	0.00
57	2,4-Dinitrotoluene	40.000	39.213	2.0	102	0.00
58	Fluorene	40.000	37.991	5.0	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.080	2.3	105	0.00
60	Diethylphthalate	40.000	37.773	5.6	102	0.00
61	4-Chlorophenyl-phenylether	40.000	37.137	7.2	102	0.00
62	4-Nitroaniline	40.000	41.063	-2.7	104	0.00
63	Azobenzene	40.000	38.275	4.3	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.745	-1.9	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.104	4.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	38.434	3.9	102	0.00
68	Hexachlorobenzene	40.000	38.884	2.8	103	0.00
69	Atrazine	40.000	39.628	0.9	102	0.00
70 C	Pentachlorophenol	40.000	41.289	-3.2	102	0.00
71	Phenanthrene	40.000	38.406	4.0	102	0.00
72	Anthracene	40.000	38.148	4.6	101	0.00
73	Carbazole	40.000	38.819	3.0	103	0.00
74	Di-n-butylphthalate	40.000	37.608	6.0	98	0.00
75 C	Fluoranthene	40.000	39.008	2.5	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	43.213	-8.0	112	0.00
78	Pyrene	40.000	37.646	5.9	103	0.00
79 S	Terphenyl-d14	80.000	72.033	10.0	102	0.00
80	Butylbenzylphthalate	40.000	38.290	4.3	103	0.00
81	Benzo(a)anthracene	40.000	38.690	3.3	108	0.00
82	3,3'-Dichlorobenzidine	40.000	39.879	0.3	110	0.00
83	Chrysene	40.000	38.680	3.3	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.331	6.7	104	0.00
85 c	Di-n-octyl phthalate	40.000	37.074	7.3	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.938	5.2	94	0.00
88	Benzo(b)fluoranthene	40.000	40.237	-0.6	105	0.00
89	Benzo(k)fluoranthene	40.000	39.533	1.2	104	0.00
90 C	Benzo(a)pyrene	40.000	39.375	1.6	101	0.00
91	Dibenzo(a,h)anthracene	40.000	37.845	5.4	95	0.00
92	Benzo(g,h,i)perylene	40.000	37.745	5.6	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021022\
Data File : BF126834.D
Acq On : 10 Feb 2022 10:35
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 10 12:45:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0