

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021115\
 Data File : BF077271.D
 Acq On : 12 Feb 2015 4:59
 Operator : TP/IZ
 Sample : G1253-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-SLDG-WT-70

Manual Integrations
 APPROVED

apatel
 2/12/2015 4:14:50 PM

Quant Time: Feb 12 11:42:05 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 10 13:30:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	69421	20.00	ng	-0.05
21) Naphthalene-d8	10.45	136	290580	20.00	ng	-0.05
38) Acenaphthene-d10	14.58	164	141291	20.00	ng	-0.05
63) Phenanthrene-d10	17.47	188	251847	20.00	ng	-0.03
75) Chrysene-d12	20.98	240	229603	20.00	ng	-0.03
86) Perylene-d12	22.61	264	211690	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	4.56	112	222061	52.59	ng	-0.03
7) Phenol-d6	7.04	99	250926	47.11	ng	0.02
23) Nitrobenzene-d5	8.85	82	205850	39.25	ng	-0.05
41) 2,4,6-Tribromophenol	16.34	330	77503	67.58	ng	-0.03
44) 2-Fluorobiphenyl	13.12	172	408044	43.95	ng	-0.05
78) Terphenyl-d14	19.72	244	402353	40.34	ng	-0.03
Target Compounds						Qvalue
15) Benzyl Alcohol	8.02	79	35646	8.27	ng	87
31) Naphthalene	10.50	128	58594	3.92	ng	98
32) Benzoic acid	10.58	122	8759m	3.83	ng	
73) Di-n-butylphthalate	18.49	149	72472	4.21	ng	98
79) Butylbenzylphthalate	20.39	149	46848	6.57	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021115\
Data File : BF077271.D
Acq On : 12 Feb 2015 4:59
Operator : TP/IZ
Sample : G1253-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-SLDG-WT-70

Manual Integrations
APPROVED

apatel
2/12/2015 4:14:50 PM

Quant Time: Feb 12 11:42:05 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 10 13:30:38 2015
Response via : Initial Calibration

