

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021122\
 Data File : BF126864.D
 Acq On : 11 Feb 2022 18:27
 Operator : CG\JU
 Sample : N1357-01DL2 250X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GROUNDWATERDL2

Quant Time: Feb 11 22:44:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.939	152	82697	20.000	ng	0.00
21) Naphthalene-d8	8.222	136	341648	20.000	ng	0.00
39) Acenaphthene-d10	9.975	164	179154	20.000	ng	-0.01
64) Phenanthrene-d10	11.469	188	333175	20.000	ng	0.00
76) Chrysene-d12	14.110	240	180986	20.000	ng	-0.01
86) Perylene-d12	15.621	264	156736	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.545	112	1544	0.287	ng	-0.01
7) Phenol-d6	6.545	99	1708	0.236	ng	-0.02
23) Nitrobenzene-d5	7.492	82	3852	0.520	ng	-0.02
42) 2,4,6-Tribromophenol	10.763	330	1059	0.551	ng	-0.01
45) 2-Fluorobiphenyl	9.292	172	7258	0.610	ng	-0.01
79) Terphenyl-d14	13.051	244	8990	0.712	ng	-0.01
Target Compounds						
33) 4-Chloroaniline	8.292	127	788613	113.046	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021122\
Data File : BF126864.D
Acq On : 11 Feb 2022 18:27
Operator : CG\JU
Sample : N1357-01DL2 250X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GROUNDWATERDL2

Quant Time: Feb 11 22:44:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

