

Data Path : Z:\HPCHEM1\BNA F\DATA\BF021215\
 Data File : BF077294.D
 Acq On : 12 Feb 2015 21:08
 Operator : TP/IZ
 Sample : G1269-01DL 5X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-059-21015DL

Manual Integrations
 APPROVED

mohammad
 2/13/2015 5:40:22 PM

Quant Time: Feb 13 04:54:24 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 13 03:50:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	12749	20.00	ng	0.01
21) Naphthalene-d8	10.41	136	57133	20.00	ng	0.00
38) Acenaphthene-d10	14.52	164	28682	20.00	ng	0.00
63) Phenanthrene-d10	17.44	188	55553	20.00	ng	0.00
75) Chrysene-d12	20.95	240	55968	20.00	ng	-0.01
86) Perylene-d12	22.60	264	47344	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	4.53	112	17005	21.91	ng	0.05
7) Phenol-d6	6.98	99	12628	12.51	ng	0.05
23) Nitrobenzene-d5	8.81	82	36480	36.47	ng	-0.01
41) 2,4,6-Tribromophenol	16.31	330	14409	61.61	ng	0.00
44) 2-Fluorobiphenyl	13.06	172	84967	44.59	ng	-0.01
78) Terphenyl-d14	19.69	244	96688	38.80	ng	-0.01
Target Compounds						
6) Aniline	6.83	93	3324	2.49	ng	Qvalue # 87
13) 1,4-Dichlorobenzene	7.54	146	7647	7.25	ng	90
14) 1,2-Dichlorobenzene	7.85	146	140034	140.15	ng	96
30) 1,2,4-Trichlorobenzene	10.33	180	57367	68.53	ng	95
31) Naphthalene	10.45	128	138238	49.02	ng	97
33) 4-Chloroaniline	10.68	127	52066m	45.77	ng	
37) 2-Methylnaphthalene	12.11	142	30854	14.59	ng	97
51) Acenaphthene	14.60	154	6723	4.17	ng	# 85
56) 2,4-Dinitrotoluene	15.22	165	2967m	5.46	ng	
57) Fluorene	15.80	166	8377	4.10	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF021215\
Data File : BF077294.D
Acq On : 12 Feb 2015 21:08
Operator : TP/IZ
Sample : G1269-01DL 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-059-21015DL

Manual Integrations
APPROVED

mohammad
2/13/2015 5:40:22 PM

Quant Time: Feb 13 04:54:24 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 13 03:50:26 2015
Response via : Initial Calibration

