


Data Path : Z:\HPCHEM1\BNA_F\Data\BF021215\
 Data File : BF077283.D
 Acq On : 12 Feb 2015 13:32
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021215.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 13 03:50:26 2015



AutoFind: Scans 1502, 1503, 1504; Background Corrected with Scan 1499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.5	11045	PASS
68	69	0.00	2	1.3	143	PASS
69	198	0.00	100	42.6	11353	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.6	14017	PASS
197	198	0.00	2	0.4	119	PASS
198	198	100	100	100.0	26623	PASS
199	198	5	9	6.4	1704	PASS
275	198	10	60	22.6	6005	PASS
365	198	1	100	2.9	763	PASS
441	442	0.01	24	14.4	3472	PASS
442	198	50	100	90.7	24145	PASS
443	442	15	24	18.6	4481	PASS