

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021216\
 Data File : BF085036.D
 Acq On : 12 Feb 2016 2:49
 Operator : SJ/IZ
 Sample : PB88279TB
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB88279TB

Quant Time: Feb 12 04:34:26 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 10 13:28:51 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.60	152	138780	20.00	ng	-0.02
21) Naphthalene-d8	7.90	136	608541	20.00	ng	-0.02
38) Acenaphthene-d10	9.64	164	288318	20.00	ng	-0.02
63) Phenanthrene-d10	11.12	188	487451	20.00	ng	-0.02
75) Chrysene-d12	13.75	240	325101	20.00	ng	-0.02
86) Perylene-d12	15.11	264	276001	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	978815	109.86	ng	0.00
7) Phenol-d6	6.26	99	1194435	108.14	ng	-0.02
23) Nitrobenzene-d5	7.18	82	740714	68.57	ng	-0.02
41) 2,4,6-Tribromophenol	10.43	330	299673	99.65	ng	-0.02
44) 2-Fluorobiphenyl	8.98	172	1569280	95.70	ng	-0.01
78) Terphenyl-d14	12.71	244	1291065	89.99	ng	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021216\
Data File : BF085036.D
Acq On : 12 Feb 2016 2:49
Operator : SJ/IZ
Sample : PB88279TB
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB88279TB

Quant Time: Feb 12 04:34:26 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 10 13:28:51 2016
Response via : Initial Calibration

