

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF021218\
 Data File : BF102910.D
 Acq On : 12 Feb 2018 10:35
 Operator : SJ/JU
 Sample : J1506-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-22(2-4)

Manual Integrations
 APPROVED

Sohil
 2/13/2018 11:25:08 AM

Quant Time: Feb 12 12:06:39 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 12:44:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	176925	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	746359	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	339594	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	566319	20.00	ng	0.00
75) Chrysene-d12	14.05	240	368598	20.00	ng	0.00
86) Perylene-d12	15.54	264	330994	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	1445442	124.07	ng	0.02
7) Phenol-d6	6.52	99	1716898	122.40	ng	0.01
23) Nitrobenzene-d5	7.45	82	1132424	105.25	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	348345	127.44	ng	0.00
44) 2-Fluorobiphenyl	9.25	172	1521817	74.36	ng	0.00
78) Terphenyl-d14	13.00	244	1264096	81.20	ng	0.00
Target Compounds						
10) Phenol	6.53	94	41527	2.762	ng	99
49) Dimethylphthalate	9.63	163	151440	5.719	ng	97
74) Fluoranthene	12.62	202	76819	2.421	ng	98
77) Pyrene	12.86	202	66240	2.292	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF021218\
Data File : BF102910.D
Acq On : 12 Feb 2018 10:35
Operator : SJ/JU
Sample : J1506-07
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-22(2-4)

Manual Integrations
APPROVED

Sohil
2/13/2018 11:25:08 AM

Quant Time: Feb 12 12:06:39 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 05 12:44:16 2018
Response via : Initial Calibration

