

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF021218\
 Data File : BF102911.D
 Acq On : 12 Feb 2018 11:02
 Operator : SJ/JU
 Sample : J1507-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-214

Manual Integrations
 APPROVED

Sohil
 2/13/2018 11:25:10 AM

Quant Time: Feb 12 12:12:05 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 12:44:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	184767	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	756254	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	312517	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	456400	20.00	ng	0.00
75) Chrysene-d12	14.05	240	336689	20.00	ng	0.00
86) Perylene-d12	15.54	264	362500	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	1171529	96.29	ng	0.02
7) Phenol-d6	6.52	99	1362577	93.01	ng	0.01
23) Nitrobenzene-d5	7.45	82	902668	82.80	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	234285	93.14	ng	0.00
44) 2-Fluorobiphenyl	9.24	172	1305372	69.31	ng	0.00
78) Terphenyl-d14	13.00	244	828776	58.28	ng	0.00
Target Compounds						
10) Phenol	6.52	94	74523	4.747	ng	85
49) Dimethylphthalate	9.63	163	145934	5.989	ng	100
70) Phenanthrene	11.43	178	99343	3.908	ng	97
74) Fluoranthene	12.62	202	177594	6.944	ng	96
77) Pyrene	12.86	202	161552	6.119	ng	98
80) Benzo(a)anthracene	14.04	228	81104	3.830	ng	100
82) Chrysene	14.07	228	94016	4.405	ng	97
83) Bis(2-ethylhexyl)phthalate	14.02	149	47985	2.967	ng	# 99
85) Indeno(1,2,3-cd)pyrene	17.04	276	42979	2.428	ng	96
87) Benzo(b)fluoranthene	15.10	252	114140m	4.857	ng	
89) Benzo(a)pyrene	15.47	252	77301	3.654	ng	96
91) Benzo(a,h,i)perylene	17.49	276	42393	2.522	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF021218\
Data File : BF102911.D
Acq On : 12 Feb 2018 11:02
Operator : SJ/JU
Sample : J1507-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-214

Manual Integrations
APPROVED

Sohil
2/13/2018 11:25:10 AM

Quant Time: Feb 12 12:12:05 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 05 12:44:16 2018
Response via : Initial Calibration

