

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021221\
 Data File : BF123199.D
 Acq On : 12 Feb 2021 16:31
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF021221

Quant Time: Feb 12 17:15:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 16:59:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	39.176	2.1	100	0.00
3	Pyridine	40.000	39.516	1.2	100	0.00
4	n-Nitrosodimethylamine	40.000	39.635	0.9	100	0.00
5 S	2-Fluorophenol	80.000	78.939	1.3	100	0.00
6	Aniline	40.000	39.867	0.3	100	0.00
7 S	Phenol-d6	80.000	78.509	1.9	101	0.00
8	2-Chlorophenol	40.000	39.463	1.3	101	0.00
9	Benzaldehyde	40.000	37.041	7.4	102	0.00
10 C	Phenol	40.000	39.393	1.5	102	0.00
11	bis(2-Chloroethyl)ether	40.000	39.302	1.7	100	0.00
12	1,3-Dichlorobenzene	40.000	39.181	2.0	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.674	0.8	102	0.00
14	1,2-Dichlorobenzene	40.000	39.433	1.4	100	0.00
15	Benzyl Alcohol	40.000	40.292	-0.7	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.251	1.9	99	0.00
17	2-Methylphenol	40.000	39.283	1.8	100	0.00
18	Hexachloroethane	40.000	40.254	-0.6	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.305	1.7	102	0.00
20	3+4-Methylphenols	40.000	39.107	2.2	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	38.573	3.6	102	0.00
23 S	Nitrobenzene-d5	80.000	79.085	1.1	101	0.00
24	Nitrobenzene	40.000	39.789	0.5	102	0.00
25	Isophorone	40.000	39.836	0.4	102	0.00
26 C	2-Nitrophenol	40.000	40.934	-2.3	102	0.00
27	2,4-Dimethylphenol	40.000	39.491	1.3	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.524	1.2	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.066	-0.2	101	0.00
30	1,2,4-Trichlorobenzene	40.000	39.318	1.7	100	0.00
31	Naphthalene	40.000	39.225	1.9	101	0.00
32	Benzoic acid	40.000	43.784	-9.5	105	0.00
33	4-Chloroaniline	40.000	39.489	1.3	102	0.00
34 C	Hexachlorobutadiene	40.000	38.924	2.7	99	0.00
35	Caprolactam	40.000	40.124	-0.3	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.774	-1.9	104	0.00
37	2-Methylnaphthalene	40.000	39.568	1.1	101	0.00
38	1-Methylnaphthalene	40.000	39.509	1.2	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.875	2.8	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.638	-4.1	97	0.00
42 S	2,4,6-Tribromophenol	80.000	80.057	-0.1	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.940	0.2	102	0.00
44	2,4,5-Trichlorophenol	40.000	39.842	0.4	103	0.00
45 S	2-Fluorobiphenyl	80.000	76.527	4.3	101	0.00
46	1,1'-Biphenyl	40.000	39.344	1.6	102	0.00
47	2-Chloronaphthalene	40.000	39.023	2.4	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021221\
 Data File : BF123199.D
 Acq On : 12 Feb 2021 16:31
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF021221

Quant Time: Feb 12 17:15:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 16:59:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.445	-1.1	103	0.00
49	Acenaphthylene	40.000	39.073	2.3	101	0.00
50	Dimethylphthalate	40.000	39.459	1.4	102	0.00
51	2,6-Dinitrotoluene	40.000	41.084	-2.7	104	0.00
52 C	Acenaphthene	40.000	39.552	1.1	102	0.00
53	3-Nitroaniline	40.000	40.255	-0.6	102	0.00
54 P	2,4-Dinitrophenol	40.000	43.010	-7.5	105	0.00
55	Dibenzofuran	40.000	39.319	1.7	102	0.00
56 P	4-Nitrophenol	40.000	41.476	-3.7	102	0.00
57	2,4-Dinitrotoluene	40.000	40.609	-1.5	103	0.00
58	Fluorene	40.000	39.117	2.2	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.730	0.7	101	0.00
60	Diethylphthalate	40.000	39.728	0.7	102	0.00
61	4-Chlorophenyl-phenylether	40.000	39.082	2.3	101	0.00
62	4-Nitroaniline	40.000	40.081	-0.2	102	0.00
63	Azobenzene	40.000	38.927	2.7	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.061	-7.7	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.348	1.6	102	0.00
67	4-Bromophenyl-phenylether	40.000	39.264	1.8	101	0.00
68	Hexachlorobenzene	40.000	39.234	1.9	102	0.00
69	Atrazine	40.000	38.455	3.9	100	0.00
70 C	Pentachlorophenol	40.000	42.414	-6.0	104	0.00
71	Phenanthrene	40.000	38.516	3.7	101	0.00
72	Anthracene	40.000	38.757	3.1	101	0.00
73	Carbazole	40.000	39.530	1.2	104	0.00
74	Di-n-butylphthalate	40.000	39.842	0.4	101	0.00
75 C	Fluoranthene	40.000	39.359	1.6	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	36.261	9.3	93	0.00
78	Pyrene	40.000	39.310	1.7	102	0.00
79 S	Terphenyl-d14	80.000	77.529	3.1	102	0.00
80	Butylbenzylphthalate	40.000	39.831	0.4	100	0.00
81	Benzo(a)anthracene	40.000	39.721	0.7	103	0.00
82	3,3'-Dichlorobenzidine	40.000	39.613	1.0	101	0.00
83	Chrysene	40.000	38.929	2.7	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.704	0.7	100	0.00
85 c	Di-n-octyl phthalate	40.000	40.248	-0.6	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.406	-1.0	104	0.00
88	Benzo(b)fluoranthene	40.000	38.651	3.4	100	0.00
89	Benzo(k)fluoranthene	40.000	40.817	-2.0	104	0.00
90 C	Benzo(a)pyrene	40.000	39.621	0.9	102	0.00
91	Dibenzo(a,h)anthracene	40.000	40.632	-1.6	104	0.00
92	Benzo(g,h,i)perylene	40.000	40.129	-0.3	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021221\
Data File : BF123199.D
Acq On : 12 Feb 2021 16:31
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF021221

Quant Time: Feb 12 17:15:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 12 16:59:49 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0