




Data Path : Z:\HPCHEM1\BNA F\DATA\BF021317\
 Data File : BF092917.D
 Acq On : 13 Feb 2017 14:17
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 13 18:19:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 13 18:04:38 2017
 Response via : Initial Calibration



TIC: BF092917.D

(69) Pentachlorophenol (C)




11.208min (-0.011) 0.00ng

response 164641

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	62.79
264.00	62.90	61.73
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF021317\
Data File : BF092917.D
Acq On : 13 Feb 2017 14:17
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 13 18:19:21 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 13 18:04:38 2017
Response via : Initial Calibration



TIC: BF092917.D

(76) Benzidine

12.739min (+0.000) 0.00ng

response 834522

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	14.20
183.00	11.50	12.36
0.00	0.00	0.00