

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021323\
 Data File : BF132417.D
 Acq On : 13 Feb 2023 19:07
 Operator : CG\JU
 Sample : 01442-02RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EB-2023-2-3RE

Quant Time: Feb 14 01:39:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 14 00:51:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.031	152	155944	20.000	ng	0.00
21) Naphthalene-d8	8.319	136	592330	20.000	ng	0.00
39) Acenaphthene-d10	10.083	164	305505	20.000	ng	0.00
64) Phenanthrene-d10	11.577	188	575016	20.000	ng	0.00
76) Chrysene-d12	14.236	240	432087	20.000	ng	0.00
86) Perylene-d12	15.801	264	452793	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.672	112	1478900	173.059	ng	0.02
7) Phenol-d6	6.654	99	1300338	122.093	ng	0.00
23) Nitrobenzene-d5	7.607	82	1603388	161.483	ng	0.00
42) 2,4,6-Tribromophenol	10.883	330	1122442	276.090	ng	0.00
45) 2-Fluorobiphenyl	9.401	172	3039836	144.048	ng	0.00
79) Terphenyl-d14	13.183	244	3879222	150.993	ng	0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021323\
Data File : BF132417.D
Acq On : 13 Feb 2023 19:07
Operator : CG\JU
Sample : 01442-02RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EB-2023-2-3RE

Quant Time: Feb 14 01:39:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 14 00:51:00 2023
Response via : Initial Calibration

