

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021323\
 Data File : BF132418.D
 Acq On : 13 Feb 2023 19:40
 Operator : CG\JU
 Sample : 01329-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-7-BORING-2-14-21

Quant Time: Feb 14 01:40:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 14 00:51:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.031	152	156211	20.000	ng	0.00
21) Naphthalene-d8	8.319	136	587236	20.000	ng	0.00
39) Acenaphthene-d10	10.084	164	308543	20.000	ng	0.00
64) Phenanthrene-d10	11.578	188	577259	20.000	ng	0.00
76) Chrysene-d12	14.236	240	444434	20.000	ng	0.00
86) Perylene-d12	15.801	264	453557	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.666	112	1255833	146.705	ng	0.02
7) Phenol-d6	6.654	99	1480897	138.808	ng	0.00
23) Nitrobenzene-d5	7.601	82	905068	91.943	ng	0.00
42) 2,4,6-Tribromophenol	10.878	330	578756	140.957	ng	0.00
45) 2-Fluorobiphenyl	9.395	172	1729383	81.143	ng	0.00
79) Terphenyl-d14	13.177	244	2020920	76.476	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021323\
Data File : BF132418.D
Acq On : 13 Feb 2023 19:40
Operator : CG\JU
Sample : 01329-06
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-7-BORING-2-14-21

Quant Time: Feb 14 01:40:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 14 00:51:00 2023
Response via : Initial Calibration

