

Data Path : Z:\HPCHEM1\BNA F\DATA\BF021416\
 Data File : BF085074.D
 Acq On : 13 Feb 2016 9:05
 Operator : SJ/IZ
 Sample : H1409-05
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SUP-B002(12-13.5)

Quant Time: Feb 15 01:37:20 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 13 00:11:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.92	152	114790	20.00	ng	0.03
21) Naphthalene-d8	7.79	136	556809	20.00	ng	0.00
38) Acenaphthene-d10	10.58	164	279229	20.00	ng	0.00
63) Phenanthrene-d10	12.99	188	531406	20.00	ng	0.00
75) Chrysene-d12	17.18	240	350567	20.00	ng	0.00
86) Perylene-d12	18.80	264	265162	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.14	112	1070575	163.08	ng	-0.02
7) Phenol-d6	5.43	99	1537265	175.70	ng	-0.01
23) Nitrobenzene-d5	6.72	82	964146	102.25	ng	0.00
41) 2,4,6-Tribromophenol	11.88	330	442849	153.55	ng	0.00
44) 2-Fluorobiphenyl	9.53	172	1930456	98.82	ng	0.00
78) Terphenyl-d14	15.62	244	1526188	89.60	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.22	163	361694	17.28	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF021416\
Data File : BF085074.D
Acq On : 13 Feb 2016 9:05
Operator : SJ/IZ
Sample : H1409-05
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SUP-B002(12-13.5)

Quant Time: Feb 15 01:37:20 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 13 00:11:18 2016
Response via : Initial Calibration

