

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
 Data File : BF126873.D
 Acq On : 14 Feb 2022 16:14
 Operator : CG\JU
 Sample : N1357-01DL3 1250X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GROUNDWATERDL3

Quant Time: Feb 15 00:59:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.940	152	103494	20.000	ng	0.00
21) Naphthalene-d8	8.222	136	431459	20.000	ng	0.00
39) Acenaphthene-d10	9.980	164	224646	20.000	ng	0.00
64) Phenanthrene-d10	11.469	188	379001	20.000	ng	0.00
76) Chrysene-d12	14.110	240	176558	20.000	ng	-0.01
86) Perylene-d12	15.621	264	180087	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.551	112	158	0.024	ng	0.00
7) Phenol-d6	6.551	99	210	0.023	ng	-0.01
23) Nitrobenzene-d5	7.492	82	838	0.090	ng	-0.02
42) 2,4,6-Tribromophenol	10.769	330	73	0.030	ng	0.00
45) 2-Fluorobiphenyl	9.292	172	1620	0.109	ng	-0.01
79) Terphenyl-d14	13.051	244	1350	0.110	ng	-0.01
Target Compounds						
33) 4-Chloroaniline	8.286	127	189197	21.476	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
Data File : BF126873.D
Acq On : 14 Feb 2022 16:14
Operator : CG\JU
Sample : N1357-01DL3 1250X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GROUNDWATERDL3

Quant Time: Feb 15 00:59:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

