

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
 Data File : BF126867.D
 Acq On : 14 Feb 2022 13:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 00:57:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	0.000	100.0#	0	-0.02
3	Pyridine	40.000	37.641	5.9	89	0.00
4	n-Nitrosodimethylamine	40.000	37.744	5.6	90	0.00
5 S	2-Fluorophenol	80.000	75.247	5.9	90	0.00
6	Aniline	40.000	37.930	5.2	89	0.00
7 S	Phenol-d6	80.000	75.651	5.4	90	0.00
8	2-Chlorophenol	40.000	38.022	4.9	90	0.00
9	Benzaldehyde	40.000	34.470	13.8	88	0.00
10 C	Phenol	40.000	37.251	6.9	89	0.00
11	bis(2-Chloroethyl)ether	40.000	37.554	6.1	89	0.00
12	1,3-Dichlorobenzene	40.000	37.466	6.3	89	0.00
13 C	1,4-Dichlorobenzene	40.000	37.315	6.7	89	0.00
14	1,2-Dichlorobenzene	40.000	38.064	4.8	90	0.00
15	Benzyl Alcohol	40.000	40.172	-0.4	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.814	10.5	84	0.00
17	2-Methylphenol	40.000	38.009	5.0	89	0.00
18	Hexachloroethane	40.000	37.945	5.1	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.406	6.5	89	0.00
20	3+4-Methylphenols	40.000	38.107	4.7	90	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	37.150	7.1	91	0.00
23 S	Nitrobenzene-d5	80.000	74.703	6.6	90	0.00
24	Nitrobenzene	40.000	37.866	5.3	91	0.00
25	Isophorone	40.000	36.901	7.7	89	0.00
26 C	2-Nitrophenol	40.000	39.082	2.3	91	0.00
27	2,4-Dimethylphenol	40.000	37.536	6.2	89	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.624	10.9	86	0.00
29 C	2,4-Dichlorophenol	40.000	37.323	6.7	91	0.00
30	1,2,4-Trichlorobenzene	40.000	36.096	9.8	88	0.00
31	Naphthalene	40.000	36.555	8.6	89	0.00
32	Benzoic acid	40.000	39.490	1.3	87	0.00
33	4-Chloroaniline	40.000	37.171	7.1	90	0.00
34 C	Hexachlorobutadiene	40.000	37.565	6.1	93	0.00
35	Caprolactam	40.000	38.576	3.6	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.089	4.8	91	0.00
37	2-Methylnaphthalene	40.000	36.470	8.8	89	0.00
38	1-Methylnaphthalene	40.000	36.753	8.1	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.089	7.3	91	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.016	-12.5	104	0.00
42 S	2,4,6-Tribromophenol	80.000	78.436	2.0	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.356	4.1	94	0.00
44	2,4,5-Trichlorophenol	40.000	37.967	5.1	92	0.00
45 S	2-Fluorobiphenyl	80.000	72.968	8.8	91	0.00
46	1,1'-Biphenyl	40.000	36.438	8.9	90	0.00
47	2-Chloronaphthalene	40.000	36.801	8.0	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
 Data File : BF126867.D
 Acq On : 14 Feb 2022 13:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 00:57:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.425	1.4	92	0.00
49	Acenaphthylene	40.000	36.817	8.0	89	0.00
50	Dimethylphthalate	40.000	37.143	7.1	91	0.00
51	2,6-Dinitrotoluene	40.000	37.855	5.4	89	0.00
52 C	Acenaphthene	40.000	37.284	6.8	90	0.00
53	3-Nitroaniline	40.000	38.560	3.6	92	0.00
54 P	2,4-Dinitrophenol	40.000	36.572	8.6	87	0.00
55	Dibenzofuran	40.000	36.921	7.7	91	0.00
56 P	4-Nitrophenol	40.000	39.791	0.5	90	0.00
57	2,4-Dinitrotoluene	40.000	39.002	2.5	92	0.00
58	Fluorene	40.000	36.854	7.9	92	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.125	4.7	93	0.00
60	Diethylphthalate	40.000	36.695	8.3	90	0.00
61	4-Chlorophenyl-phenylether	40.000	36.912	7.7	92	0.00
62	4-Nitroaniline	40.000	39.725	0.7	92	0.00
63	Azobenzene	40.000	36.642	8.4	89	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.026	4.9	88	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.471	8.8	91	0.00
67	4-Bromophenyl-phenylether	40.000	36.727	8.2	91	0.00
68	Hexachlorobenzene	40.000	37.108	7.2	92	0.00
69	Atrazine	40.000	38.050	4.9	91	0.00
70 C	Pentachlorophenol	40.000	39.662	0.8	91	0.00
71	Phenanthrene	40.000	36.471	8.8	91	0.00
72	Anthracene	40.000	36.233	9.4	89	0.00
73	Carbazole	40.000	36.774	8.1	91	0.00
74	Di-n-butylphthalate	40.000	36.194	9.5	88	0.00
75 C	Fluoranthene	40.000	37.023	7.4	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	40.219	-0.5	95	0.00
78	Pyrene	40.000	36.582	8.5	92	0.00
79 S	Terphenyl-d14	80.000	72.126	9.8	93	0.00
80	Butylbenzylphthalate	40.000	37.175	7.1	91	0.00
81	Benzo(a)anthracene	40.000	36.928	7.7	94	0.00
82	3,3'-Dichlorobenzidine	40.000	38.010	5.0	96	0.00
83	Chrysene	40.000	36.289	9.3	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	34.989	12.5	89	0.00
85 c	Di-n-octyl phthalate	40.000	34.085	14.8	86	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.510	6.2	86	0.00
88	Benzo(b)fluoranthene	40.000	38.931	2.7	94	0.00
89	Benzo(k)fluoranthene	40.000	36.537	8.7	89	0.00
90 C	Benzo(a)pyrene	40.000	37.665	5.8	89	0.00
91	Dibenzo(a,h)anthracene	40.000	37.482	6.3	87	0.00
92	Benzo(g,h,i)perylene	40.000	37.385	6.5	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
Data File : BF126867.D
Acq On : 14 Feb 2022 13:21
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 15 00:57:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0