

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
 Data File : BF126867.D
 Acq On : 14 Feb 2022 13:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 00:57:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.000	0.600	0.0	0#	-0.02
3	Pyridine	1.648	1.550	5.9	89	0.00
4	n-Nitrosodimethylamine	0.774	0.730	5.7	90	0.00
5 S	2-Fluorophenol	1.299	1.222	5.9	90	0.00
6	Aniline	2.112	2.003	5.2	89	0.00
7 S	Phenol-d6	1.754	1.658	5.5	90	0.00
8	2-Chlorophenol	1.383	1.314	5.0	90	0.00
9	Benzaldehyde	1.066	0.919	13.8	88	0.00
10 C	Phenol	1.768	1.646	6.9	89	0.00
11	bis(2-Chloroethyl)ether	1.407	1.321	6.1	89	0.00
12	1,3-Dichlorobenzene	1.504	1.409	6.3	89	0.00
13 C	1,4-Dichlorobenzene	1.520	1.418	6.7	89	0.00
14	1,2-Dichlorobenzene	1.433	1.363	4.9	90	0.00
15	Benzyl Alcohol	1.386	1.392	-0.4	93	0.00
16	2,2'-oxybis(1-Chloropropane	2.300	2.060	10.4	84	0.00
17	2-Methylphenol	1.210	1.149	5.0	89	0.00
18	Hexachloroethane	0.576	0.546	5.2	90	0.00
19 P	n-Nitroso-di-n-propylamine	1.110	1.038	6.5	89	0.00
20	3+4-Methylphenols	1.552	1.478	4.8	90	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.517	0.481	7.0	91	0.00
23 S	Nitrobenzene-d5	0.433	0.405	6.5	90	0.00
24	Nitrobenzene	0.410	0.388	5.4	91	0.00
25	Isophorone	0.726	0.670	7.7	89	0.00
26 C	2-Nitrophenol	0.172	0.168	2.3	91	0.00
27	2,4-Dimethylphenol	0.290	0.272	6.2	89	0.00
28	bis(2-Chloroethoxy)methane	0.451	0.402	10.9	86	0.00
29 C	2,4-Dichlorophenol	0.272	0.254	6.6	91	0.00
30	1,2,4-Trichlorobenzene	0.301	0.272	9.6	88	0.00
31	Naphthalene	1.037	0.948	8.6	89	0.00
32	Benzoic acid	0.204	0.202	1.0	87	0.00
33	4-Chloroaniline	0.408	0.379	7.1	90	0.00
34 C	Hexachlorobutadiene	0.171	0.161	5.8	93	0.00
35	Caprolactam	0.093	0.090	3.2	90	0.00
36 C	4-Chloro-3-methylphenol	0.343	0.326	5.0	91	0.00
37	2-Methylnaphthalene	0.663	0.604	8.9	89	0.00
38	1-Methylnaphthalene	0.641	0.589	8.1	90	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	0.522	0.484	7.3	91	0.00
41 P	Hexachlorocyclopentadiene	0.294	0.331	-12.6	104	0.00
42 S	2,4,6-Tribromophenol	0.214	0.210	1.9	95	0.00
43 C	2,4,6-Trichlorophenol	0.377	0.361	4.2	94	0.00
44	2,4,5-Trichlorophenol	0.399	0.378	5.3	92	0.00
45 S	2-Fluorobiphenyl	1.329	1.212	8.8	91	0.00
46	1,1'-Biphenyl	1.586	1.445	8.9	90	0.00
47	2-Chloronaphthalene	1.172	1.078	8.0	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
 Data File : BF126867.D
 Acq On : 14 Feb 2022 13:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 00:57:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.455	0.448	1.5	92	0.00
49	Acenaphthylene	1.909	1.757	8.0	89	0.00
50	Dimethylphthalate	1.415	1.314	7.1	91	0.00
51	2,6-Dinitrotoluene	0.287	0.271	5.6	89	0.00
52 C	Acenaphthene	1.221	1.138	6.8	90	0.00
53	3-Nitroaniline	0.350	0.337	3.7	92	0.00
54 P	2,4-Dinitrophenol	0.142	0.137	3.5	87	0.00
55	Dibenzofuran	1.681	1.552	7.7	91	0.00
56 P	4-Nitrophenol	0.251	0.250	0.4	90	0.00
57	2,4-Dinitrotoluene	0.379	0.369	2.6	92	0.00
58	Fluorene	1.289	1.188	7.8	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.312	0.298	4.5	93	0.00
60	Diethylphthalate	1.460	1.339	8.3	90	0.00
61	4-Chlorophenyl-phenylether	0.604	0.558	7.6	92	0.00
62	4-Nitroaniline	0.329	0.326	0.9	92	0.00
63	Azobenzene	1.662	1.522	8.4	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	0.114	0.109	4.4	88	0.00
66 c	n-Nitrosodiphenylamine	0.654	0.597	8.7	91	0.00
67	4-Bromophenyl-phenylether	0.217	0.200	7.8	91	0.00
68	Hexachlorobenzene	0.234	0.217	7.3	92	0.00
69	Atrazine	0.200	0.190	5.0	91	0.00
70 C	Pentachlorophenol	0.129	0.128	0.8	91	0.00
71	Phenanthrene	1.108	1.010	8.8	91	0.00
72	Anthracene	1.107	1.003	9.4	89	0.00
73	Carbazole	1.008	0.927	8.0	91	0.00
74	Di-n-butylphthalate	1.299	1.176	9.5	88	0.00
75 C	Fluoranthene	1.037	0.959	7.5	92	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.526	0.529	-0.6	95	0.00
78	Pyrene	1.815	1.660	8.5	92	0.00
79 S	Terphenyl-d14	1.395	1.258	9.8	93	0.00
80	Butylbenzylphthalate	0.831	0.773	7.0	91	0.00
81	Benzo(a)anthracene	1.355	1.251	7.7	94	0.00
82	3,3'-Dichlorobenzidine	0.404	0.383	5.2	96	0.00
83	Chrysene	1.272	1.154	9.3	94	0.00
84	Bis(2-ethylhexyl)phthalate	1.078	0.943	12.5	89	0.00
85 c	Di-n-octyl phthalate	1.480	1.261	14.8	86	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	1.426	1.337	6.2	86	0.00
88	Benzo(b)fluoranthene	1.278	1.243	2.7	94	0.00
89	Benzo(k)fluoranthene	1.319	1.205	8.6	89	0.00
90 C	Benzo(a)pyrene	1.043	0.982	5.8	89	0.00
91	Dibenzo(a,h)anthracene	1.184	1.110	6.2	87	0.00
92	Benzo(g,h,i)perylene	1.178	1.101	6.5	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
Data File : BF126867.D
Acq On : 14 Feb 2022 13:21
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 15 00:57:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0