

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
 Data File : BF126874.D
 Acq On : 14 Feb 2022 16:42
 Operator : CG\JU
 Sample : N1436-02DL2 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 289DL2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 02/15/2022
 Supervised By :mohammad ahmed 02/15/2022

Quant Time: Feb 15 01:00:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.939	152	108042	20.000	ng	0.00
21) Naphthalene-d8	8.222	136	453919	20.000	ng	0.00
39) Acenaphthene-d10	9.980	164	231614	20.000	ng	0.00
64) Phenanthrene-d10	11.469	188	383796	20.000	ng	0.00
76) Chrysene-d12	14.110	240	184965	20.000	ng	-0.01
86) Perylene-d12	15.621	264	201857	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.545	112	78533	11.191	ng	-0.01
7) Phenol-d6	6.545	99	106462	11.238	ng	-0.02
23) Nitrobenzene-d5	7.492	82	72031	7.322	ng	-0.02
42) 2,4,6-Tribromophenol	10.763	330	23362	9.409	ng	-0.01
45) 2-Fluorobiphenyl	9.292	172	132732	8.626	ng	-0.01
79) Terphenyl-d14	13.051	244	106344	8.242	ng	-0.01
Target Compounds					Qvalue	
31) Naphthalene	8.245	128	281183	11.944	ng	98
37) 2-Methylnaphthalene	8.933	142	152978	10.169	ng	98
38) 1-Methylnaphthalene	9.033	142	92285	6.340	ng	98
46) 1,1'-Biphenyl	9.392	154	44777	2.437	ng	97
52) Acenaphthene	10.010	154	133222	9.422	ng	96
55) Dibenzofuran	10.180	168	186943	9.603	ng	98
58) Fluorene	10.527	166	188147	12.600	ng	98
71) Phenanthrene	11.498	178	826660	38.873	ng	99
72) Anthracene	11.545	178	133208	6.269	ng	99
73) Carbazole	11.698	167	46896	2.425	ng	99
75) Fluoranthene	12.686	202	384431	19.326	ng	98
78) Pyrene	12.916	202	256738	15.293	ng	100
81) Benzo(a)anthracene	14.104	228	67624	5.397	ng	98
83) Chrysene	14.139	228	56743	4.823	ng	97
88) Benzo(b)fluoranthene	15.174	252	49966m	3.875	ng	
90) Benzo(a)pyrene	15.551	252	31834	3.025	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021422\
Data File : BF126874.D
Acq On : 14 Feb 2022 16:42
Operator : CG\JU
Sample : N1436-02DL2 10X
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
289DL2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022


Quant Time: Feb 15 01:00:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.939 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2




Tgt Ion:152 Resp: 10804
Ion Ratio Lower Upper
152 100
150 156.9 124.3 186.5
115 67.7 50.5 75.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022




#5
2-Fluorophenol
Concen: 11.191 ng
RT: 5.545 min Scan# 554
Delta R.T. -0.012 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Tgt Ion:112 Resp: 78533
Ion Ratio Lower Upper
112 100
64 67.6 54.3 81.5
63 34.5 27.5 41.3





#7
Phenol-d6
Concen: 11.238 ng
RT: 6.545 min Scan# 71
Delta R.T. -0.018 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2




Tgt Ion: 99 Resp: 106462
Ion Ratio Lower Upper
99 100
42 19.6 15.8 23.6
71 37.5 30.6 45.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.222 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Tgt Ion:136 Resp: 453919
Ion Ratio Lower Upper
136 100
137 11.0 8.6 12.8
54 9.2 8.0 12.0
68 6.1 5.0 7.6





#23
Nitrobenzene-d5
Concen: 7.322 ng
RT: 7.492 min Scan# 81
Delta R.T. -0.018 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2



Tgt Ion: 82 Resp: 7203
Ion Ratio Lower Upper
82 100
128 34.8 31.6 47.4
54 55.9 43.7 65.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#31
Naphthalene
Concen: 11.944 ng
RT: 8.245 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42

Tgt Ion:128 Resp: 281183
Ion Ratio Lower Upper
128 100
129 10.5 8.8 13.2
127 12.3 10.6 15.8









#37
2-Methylnaphthalene
Concen: 10.169 ng
RT: 8.933 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42

Instrument :
BNA_F
ClientSampleId :
289DL2

Tgt Ion:142 Resp: 152978
Ion Ratio Lower Upper
142 100
141 89.1 69.7 104.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#38
1-Methylnaphthalene
Concen: 6.340 ng
RT: 9.033 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42


Tgt Ion:142 Resp: 92285
Ion Ratio Lower Upper
142 100
141 92.2 72.3 108.5





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.980 min Scan# 1173
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2




Tgt Ion:164 Resp: 231614
Ion Ratio Lower Upper
164 100
162 101.2 80.9 121.3
160 45.8 36.0 54.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022




#42
2,4,6-Tribromophenol
Concen: 9.409 ng
RT: 10.763 min Scan# 1441
Delta R.T. -0.012 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42



Tgt Ion:330 Resp: 23362
Ion Ratio Lower Upper
330 100
332 97.0 78.2 117.4
141 39.7 31.2 46.8





#45

2-Fluorobiphenyl

Concen: 8.626 ng

RT: 9.292 min Scan# 1191

Delta R.T. -0.012 min

Lab File: BF126874.D

Acq: 14 Feb 2022 16:42

Instrument :

BNA_F

ClientSampleId :

289DL2

Tgt Ion:172 Resp: 13273

Ion Ratio Lower Upper

172 100





171 33.7 28.5 42.7

170 23.0 18.7 28.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#46

1,1'-Biphenyl

Concen: 2.437 ng

RT: 9.392 min Scan# 1208

Delta R.T. -0.012 min

Lab File: BF126874.D

Acq: 14 Feb 2022 16:42




Tgt Ion:154 Resp: 44777


Ion Ratio Lower Upper

154 100

153 38.1 20.6 60.6

76 14.8 0.0 34.0









#52
Acenaphthene
Concen: 9.422 ng
RT: 10.010 min Scan# 1179
Delta R.T. -0.012 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42

Instrument :
BNA_F
ClientSampleId :
289DL2

Tgt Ion:154 Resp: 13322
Ion Ratio Lower Upper
154 100
153 113.3 88.5 132.7
152 56.3 41.8 62.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#55
Dibenzofuran
Concen: 9.603 ng
RT: 10.180 min Scan# 1342
Delta R.T. -0.012 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42


Tgt Ion:168 Resp: 186943
Ion Ratio Lower Upper
168 100
139 45.2 35.4 53.0
169 14.3 10.6 16.0





#58
Fluorene
Concen: 12.600 ng
RT: 10.527 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2




Tgt Ion:166 Resp: 18814
Ion Ratio Lower Upper
166 100
165 96.4 78.8 118.2
167 13.8 10.6 16.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.469 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Tgt Ion:188 Resp: 383796
Ion Ratio Lower Upper
188 100
94 11.2 8.1 12.1
80 13.0 9.5 14.3





#71
Phenanthrene
Concen: 38.873 ng
RT: 11.498 min Scan# 11574
Delta R.T. -0.000 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2




Tgt Ion:178 Resp: 826660
Ion Ratio Lower Upper
178 100
176 18.9 15.7 23.5
179 14.9 12.1 18.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022




#72
Anthracene
Concen: 6.269 ng
RT: 11.545 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Tgt Ion:178 Resp: 133208
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 15.4 12.5 18.7





#73
 Carbazole
 Concen: 2.425 ng
 RT: 11.698 min Scan# 11698
 Delta R.T. -0.006 min
 Lab File: BF126874.D
 Acq: 14 Feb 2022 16:42




Instrument :
 BNA_F
 ClientSampleId :
 289DL2




Tgt Ion:167 Resp: 46890
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.4 26.0
 139 14.3 12.2 18.4

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
 Supervised By :mohammad ahmed 02/15/2022




#75
 Fluoranthene
 Concen: 19.326 ng
 RT: 12.686 min Scan# 1768
 Delta R.T. -0.006 min
 Lab File: BF126874.D
 Acq: 14 Feb 2022 16:42




Tgt Ion:202 Resp: 384431
 Ion Ratio Lower Upper
 202 100
 101 12.1 0.0 32.5
 203 16.5 0.0 37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.110 min Scan# 20
Delta R.T. -0.012 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2



Tgt Ion:240 Resp: 18496
Ion Ratio Lower Upper
240 100
120 13.9 9.8 14.6
236 25.6 20.5 30.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#78
Pyrene
Concen: 15.293 ng
RT: 12.916 min Scan# 1807
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42


Tgt Ion:202 Resp: 256738
Ion Ratio Lower Upper
202 100
200 20.6 16.8 25.2
203 17.4 13.9 20.9





#79
 Terphenyl-d14
 Concen: 8.242 ng
 RT: 13.051 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BF126874.D
 Acq: 14 Feb 2022 16:42




Instrument :
 BNA_F
 ClientSampleId :
 289DL2




Tgt Ion:244 Resp: 106344
 Ion Ratio Lower Upper
 244 100
 212 6.4 5.6 8.4
 122 10.8 9.0 13.6

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 02/15/2022
 Supervised By :mohammad ahmed 02/15/2022




#81
 Benzo(a)anthracene
 Concen: 5.397 ng
 RT: 14.104 min Scan# 2009
 Delta R.T. -0.006 min
 Lab File: BF126874.D
 Acq: 14 Feb 2022 16:42



Tgt Ion:228 Resp: 67624
 Ion Ratio Lower Upper
 228 100
 226 27.9 21.4 32.2
 229 20.9 16.0 24.0









#83
Chrysene
Concen: 4.823 ng
RT: 14.139 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42

Instrument :
BNA_F
ClientSampleId :
289DL2

Tgt Ion:228 Resp: 5674
Ion Ratio Lower Upper
228 100
226 29.0 24.4 36.6
229 22.3 16.4 24.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.621 min Scan# 2267
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42


Tgt Ion:264 Resp: 201857
Ion Ratio Lower Upper
264 100
260 22.6 19.6 29.4
265 21.8 15.9 23.9





#88
Benzo(b)fluoranthene
Concen: 3.875 ng m
RT: 15.174 min Scan# 2191
Delta R.T. -0.006 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42




Instrument :
BNA_F
ClientSampleId :
289DL2




Tgt Ion:252 Resp: 49960
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.2
125 14.1 8.9 13.3

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 02/15/2022
Supervised By :mohammad ahmed 02/15/2022



#90
Benzo(a)pyrene
Concen: 3.025 ng
RT: 15.551 min Scan# 2255
Delta R.T. -0.012 min
Lab File: BF126874.D
Acq: 14 Feb 2022 16:42



Tgt Ion:252 Resp: 31834
Ion Ratio Lower Upper
252 100
253 23.1 16.4 24.6
125 14.6 9.1 13.7

