

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021425\
 Data File : BF141628.D
 Acq On : 14 Feb 2025 09:45
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 14 11:12:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 16:58:59 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.513	0.505	1.6	113	0.00
3	Pyridine	1.222	1.171	4.2	107	-0.01
4	n-Nitrosodimethylamine	0.809	0.747	7.7	106	-0.02
5 S	2-Fluorophenol	1.276	1.279	-0.2	115	0.00
6	Aniline	1.513	1.166	22.9	88	0.00
7 S	Phenol-d6	1.612	1.573	2.4	113	0.00
8	2-Chlorophenol	1.378	1.380	-0.1	116	0.00
9	Benzaldehyde	0.857	1.045	-21.9	147	0.00
10 C	Phenol	1.678	1.622	3.3	112	-0.01
11	bis(2-Chloroethyl)ether	1.261	1.255	0.5	115	0.00
12	1,3-Dichlorobenzene	1.455	1.470	-1.0	118	0.00
13 C	1,4-Dichlorobenzene	1.487	1.488	-0.1	117	0.00
14	1,2-Dichlorobenzene	1.380	1.389	-0.7	117	0.00
15	Benzyl Alcohol	1.236	1.167	5.6	107	0.00
16	2,2'-oxybis(1-Chloropropane	2.158	2.028	6.0	109	0.00
17	2-Methylphenol	1.079	1.058	1.9	114	0.00
18	Hexachloroethane	0.550	0.557	-1.3	117	0.00
19 P	n-Nitroso-di-n-propylamine	0.964	0.946	1.9	113	-0.01
20	3+4-Methylphenols	1.371	1.341	2.2	113	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
22	Acetophenone	0.498	0.483	3.0	114	-0.01
23 S	Nitrobenzene-d5	0.383	0.368	3.9	111	0.00
24	Nitrobenzene	0.374	0.362	3.2	113	0.00
25	Isophorone	0.611	0.597	2.3	114	-0.01
26 C	2-Nitrophenol	0.186	0.187	-0.5	113	0.00
27	2,4-Dimethylphenol	0.217	0.219	-0.9	114	0.00
28	bis(2-Chloroethoxy)methane	0.392	0.392	0.0	116	0.00
29 C	2,4-Dichlorophenol	0.294	0.300	-2.0	118	0.00
30	1,2,4-Trichlorobenzene	0.320	0.317	0.9	116	0.00
31	Naphthalene	1.041	1.025	1.5	115	0.00
32	Benzoic acid	0.226	0.199	11.9	101	-0.02
33	4-Chloroaniline	0.363	0.315	13.2	101	-0.01
34 C	Hexachlorobutadiene	0.192	0.192	0.0	116	0.00
35	Caprolactam	0.089	0.089	0.0	112	-0.02
36 C	4-Chloro-3-methylphenol	0.325	0.318	2.2	112	0.00
37	2-Methylnaphthalene	0.677	0.670	1.0	116	0.00
38	1-Methylnaphthalene	0.656	0.643	2.0	114	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	115	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.579	0.583	-0.7	115	0.00
41 P	Hexachlorocyclopentadiene	0.192	0.170	11.5	96	0.00
42 S	2,4,6-Tribromophenol	0.201	0.199	1.0	114	-0.01
43 C	2,4,6-Trichlorophenol	0.374	0.370	1.1	113	0.00
44	2,4,5-Trichlorophenol	0.405	0.398	1.7	112	0.00
45 S	2-Fluorobiphenyl	1.298	1.280	1.4	116	0.00
46	1,1'-Biphenyl	1.551	1.549	0.1	115	-0.01
47	2-Chloronaphthalene	1.147	1.134	1.1	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021425\
 Data File : BF141628.D
 Acq On : 14 Feb 2025 09:45
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 14 11:12:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 16:58:59 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.385	0.368	4.4	109	0.00
49	Acenaphthylene	1.705	1.663	2.5	113	0.00
50	Dimethylphthalate	1.358	1.359	-0.1	116	-0.01
51	2,6-Dinitrotoluene	0.297	0.297	0.0	114	0.00
52 C	Acenaphthene	1.147	1.104	3.7	110	-0.01
53	3-Nitroaniline	0.319	0.305	4.4	110	0.00
54 P	2,4-Dinitrophenol	0.176	0.166	5.7	108	0.00
55	Dibenzofuran	1.683	1.658	1.5	114	-0.01
56 P	4-Nitrophenol	0.237	0.203	14.3	94	0.00
57	2,4-Dinitrotoluene	0.395	0.392	0.8	113	0.00
58	Fluorene	1.301	1.276	1.9	114	0.00
59	2,3,4,6-Tetrachlorophenol	0.349	0.339	2.9	112	-0.01
60	Diethylphthalate	1.336	1.344	-0.6	116	-0.01
61	4-Chlorophenyl-phenylether	0.626	0.627	-0.2	116	-0.01
62	4-Nitroaniline	0.324	0.277	14.5	95	-0.02
63	Azobenzene	1.328	1.316	0.9	114	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	0.139	0.137	1.4	109	-0.01
66 c	n-Nitrosodiphenylamine	0.640	0.645	-0.8	116	-0.01
67	4-Bromophenyl-phenylether	0.224	0.232	-3.6	118	0.00
68	Hexachlorobenzene	0.230	0.233	-1.3	116	0.00
69	Atrazine	0.192	0.156	18.8	95	-0.01
70 C	Pentachlorophenol	0.147	0.134	8.8	98	0.00
71	Phenanthrene	1.101	1.083	1.6	113	-0.01
72	Anthracene	1.107	1.078	2.6	112	-0.01
73	Carbazole	0.996	0.950	4.6	109	0.00
74	Di-n-butylphthalate	1.150	1.176	-2.3	116	-0.01
75 C	Fluoranthene	1.167	1.112	4.7	109	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzydine	0.441	0.168	61.9#	33#	0.00
78	Pyrene	1.703	1.813	-6.5	107	-0.01
79 S	Terphenyl-d14	1.185	1.238	-4.5	107	0.00
80	Butylbenzylphthalate	0.590	0.651	-10.3	110	-0.01
81	Benzo(a)anthracene	1.329	1.312	1.3	101	0.00
82	3,3'-Dichlorobenzidine	0.381	0.329	13.6	90	0.00
83	Chrysene	1.228	1.199	2.4	103	-0.01
84	Bis(2-ethylhexyl)phthalate	0.665	0.765	-15.0	115	-0.01
85 c	Di-n-octyl phthalate	0.949	1.071	-12.9	114	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.236	1.441	-16.6	113	-0.01
88	Benzo(b)fluoranthene	1.343	1.259	6.3	96	0.00
89	Benzo(k)fluoranthene	1.147	1.195	-4.2	106	0.00
90 C	Benzo(a)pyrene	1.087	1.090	-0.3	102	0.00
91	Dibenzo(a,h)anthracene	1.001	1.169	-16.8	113	-0.02
92	Benzo(g,h,i)perylene	1.048	1.241	-18.4	114	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021425\
Data File : BF141628.D
Acq On : 14 Feb 2025 09:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 14 11:12:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 06 16:58:59 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0