

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021517\
 Data File : BF092966.D
 Acq On : 15 Feb 2017 22:15
 Operator : SJ/MA
 Sample : I1816-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 H-1

Manual Integrations
 APPROVED

mohammad
 2/16/2017 11:37:54 AM

Quant Time: Feb 16 02:32:25 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 13 18:04:38 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	165788	20.00	ng	-0.01
21) Naphthalene-d8	8.16	136	664854	20.00	ng	-0.01
38) Acenaphthene-d10	9.90	164	361210	20.00	ng	-0.02
63) Phenanthrene-d10	11.39	188	581466	20.00	ng	-0.01
75) Chrysene-d12	14.02	240	378558	20.00	ng	-0.03
86) Perylene-d12	15.46	264	329230	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1492269	154.42	ng	0.00
7) Phenol-d6	6.52	99	2001998	161.01	ng	-0.01
23) Nitrobenzene-d5	7.45	82	1267775	106.19	ng	-0.01
41) 2,4,6-Tribromophenol	10.70	330	393018	150.44	ng	-0.01
44) 2-Fluorobiphenyl	9.24	172	1971455	98.88	ng	0.00
78) Terphenyl-d14	12.98	244	1495458	92.82	ng	-0.01
Target Compounds						
10) Phenol	6.53	94	32952	2.27	ng	Qvalue # 60
49) Dimethylphthalate	9.63	163	207666	8.54	ng	100
74) Fluoranthene	12.60	202	153487	4.47	ng	92
77) Pyrene	12.83	202	122031	4.31	ng	97
80) Benzo(a)anthracene	14.01	228	48837	2.11	ng	99
82) Chrysene	14.04	228	44116	2.04	ng	96
87) Benzo(b)fluoranthene	15.05	252	58193m	2.69	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021517\
Data File : BF092966.D
Acq On : 15 Feb 2017 22:15
Operator : SJ/MA
Sample : I1816-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
H-1

Manual Integrations
APPROVED

mohammad
2/16/2017 11:37:54 AM

Quant Time: Feb 16 02:32:25 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 13 18:04:38 2017
Response via : Initial Calibration

