

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021517\
 Data File : BF092969.D
 Acq On : 15 Feb 2017 23:37
 Operator : SJ/MA
 Sample : I1859-11
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 280385-COMP-0-2FT

Manual Integrations
 APPROVED

mohammad
 2/16/2017 11:38:04 AM

Quant Time: Feb 16 02:41:46 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 13 18:04:38 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	154550	20.00	ng	-0.01
21) Naphthalene-d8	8.16	136	605243	20.00	ng	-0.01
38) Acenaphthene-d10	9.92	164	317478	20.00	ng	-0.01
63) Phenanthrene-d10	11.39	188	489236	20.00	ng	-0.01
75) Chrysene-d12	14.02	240	353541	20.00	ng	-0.03
86) Perylene-d12	15.46	264	311648	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	968625	107.52	ng	-0.01
7) Phenol-d6	6.51	99	1209010	104.31	ng	-0.02
23) Nitrobenzene-d5	7.44	82	814619	74.95	ng	-0.02
41) 2,4,6-Tribromophenol	10.70	330	250980	109.30	ng	-0.01
44) 2-Fluorobiphenyl	9.23	172	1243997	70.99	ng	-0.01
78) Terphenyl-d14	12.98	244	1052339	69.94	ng	-0.01
Target Compounds						
10) Phenol	6.52	94	32514	2.40	ng	# 75
37) 2-Methylnaphthalene	8.88	142	103691	5.26	ng	97
49) Dimethylphthalate	9.63	163	229739	10.74	ng	99
57) Fluorene	10.45	166	46568	2.44	ng	# 90
70) Phenanthrene	11.41	178	111818m	3.99	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF021517\
Data File : BF092969.D
Acq On : 15 Feb 2017 23:37
Operator : SJ/MA
Sample : I1859-11
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
280385-COMP-0-2FT

Manual Integrations
APPROVED

mohammad
2/16/2017 11:38:04 AM

Quant Time: Feb 16 02:41:46 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 13 18:04:38 2017
Response via : Initial Calibration

