

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021521\
 Data File : BF123202.D
 Acq On : 15 Feb 2021 13:16
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 14:00:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 16:59:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	39.086	2.3	105	0.00
3	Pyridine	40.000	39.765	0.6	106	0.00
4	n-Nitrosodimethylamine	40.000	39.966	0.1	106	0.00
5 S	2-Fluorophenol	80.000	78.847	1.4	105	0.00
6	Aniline	40.000	39.419	1.5	104	0.00
7 S	Phenol-d6	80.000	77.682	2.9	105	0.00
8	2-Chlorophenol	40.000	39.041	2.4	105	0.00
9	Benzaldehyde	40.000	36.908	7.7	107	0.00
10 C	Phenol	40.000	39.022	2.4	106	0.00
11	bis(2-Chloroethyl)ether	40.000	39.573	1.1	106	0.00
12	1,3-Dichlorobenzene	40.000	39.424	1.4	105	0.00
13 C	1,4-Dichlorobenzene	40.000	39.054	2.4	105	0.00
14	1,2-Dichlorobenzene	40.000	39.221	1.9	105	0.00
15	Benzyl Alcohol	40.000	40.181	-0.5	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.611	1.0	105	0.00
17	2-Methylphenol	40.000	39.221	1.9	105	0.00
18	Hexachloroethane	40.000	40.253	-0.6	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.110	2.2	107	0.00
20	3+4-Methylphenols	40.000	38.574	3.6	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	38.732	3.2	106	0.00
23 S	Nitrobenzene-d5	80.000	80.094	-0.1	106	0.00
24	Nitrobenzene	40.000	39.871	0.3	106	0.00
25	Isophorone	40.000	39.978	0.1	106	0.00
26 C	2-Nitrophenol	40.000	40.357	-0.9	105	0.00
27	2,4-Dimethylphenol	40.000	39.220	2.0	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.897	0.3	105	0.00
29 C	2,4-Dichlorophenol	40.000	39.613	1.0	104	0.00
30	1,2,4-Trichlorobenzene	40.000	39.739	0.7	105	0.00
31	Naphthalene	40.000	39.140	2.1	105	0.00
32	Benzoic acid	40.000	38.804	3.0	97	0.00
33	4-Chloroaniline	40.000	38.980	2.6	105	0.00
34 C	Hexachlorobutadiene	40.000	39.668	0.8	105	0.00
35	Caprolactam	40.000	39.468	1.3	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.197	-0.5	107	0.00
37	2-Methylnaphthalene	40.000	39.160	2.1	104	0.00
38	1-Methylnaphthalene	40.000	39.578	1.1	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.725	3.2	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.117	-2.8	100	0.00
42 S	2,4,6-Tribromophenol	80.000	77.536	3.1	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.235	-0.6	107	0.00
44	2,4,5-Trichlorophenol	40.000	38.904	2.7	104	0.00
45 S	2-Fluorobiphenyl	80.000	77.341	3.3	107	0.00
46	1,1'-Biphenyl	40.000	38.457	3.9	104	0.00
47	2-Chloronaphthalene	40.000	38.784	3.0	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021521\
 Data File : BF123202.D
 Acq On : 15 Feb 2021 13:16
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 14:00:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 12 16:59:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.136	2.2	104	0.00
49	Acenaphthylene	40.000	38.870	2.8	104	0.00
50	Dimethylphthalate	40.000	39.515	1.2	106	0.00
51	2,6-Dinitrotoluene	40.000	39.702	0.7	105	0.00
52 C	Acenaphthene	40.000	38.770	3.1	103	0.00
53	3-Nitroaniline	40.000	38.741	3.1	102	0.00
54 P	2,4-Dinitrophenol	40.000	38.210	4.5	97	0.00
55	Dibenzofuran	40.000	38.764	3.1	105	0.00
56 P	4-Nitrophenol	40.000	40.231	-0.6	103	0.00
57	2,4-Dinitrotoluene	40.000	40.306	-0.8	106	0.00
58	Fluorene	40.000	38.649	3.4	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.962	2.6	103	0.00
60	Diethylphthalate	40.000	39.360	1.6	105	0.00
61	4-Chlorophenyl-phenylether	40.000	39.252	1.9	105	0.00
62	4-Nitroaniline	40.000	39.595	1.0	105	0.00
63	Azobenzene	40.000	38.940	2.7	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.690	-1.7	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.909	2.7	104	0.00
67	4-Bromophenyl-phenylether	40.000	39.656	0.9	105	0.00
68	Hexachlorobenzene	40.000	39.247	1.9	106	0.00
69	Atrazine	40.000	38.767	3.1	103	0.00
70 C	Pentachlorophenol	40.000	40.392	-1.0	102	0.00
71	Phenanthrene	40.000	38.626	3.4	104	0.00
72	Anthracene	40.000	38.847	2.9	105	0.00
73	Carbazole	40.000	38.655	3.4	104	0.00
74	Di-n-butylphthalate	40.000	39.891	0.3	104	0.00
75 C	Fluoranthene	40.000	39.728	0.7	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzidine	40.000	37.875	5.3	98	0.00
78	Pyrene	40.000	39.651	0.9	104	0.00
79 S	Terphenyl-d14	80.000	78.357	2.1	105	0.00
80	Butylbenzylphthalate	40.000	40.957	-2.4	105	0.00
81	Benzo(a)anthracene	40.000	39.684	0.8	104	0.00
82	3,3'-Dichlorobenzidine	40.000	40.256	-0.6	104	0.00
83	Chrysene	40.000	39.407	1.5	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.406	-1.0	103	0.00
85 c	Di-n-octyl phthalate	40.000	40.845	-2.1	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.087	2.3	100	0.00
88	Benzo(b)fluoranthene	40.000	38.901	2.7	101	0.00
89	Benzo(k)fluoranthene	40.000	40.808	-2.0	104	0.00
90 C	Benzo(a)pyrene	40.000	39.250	1.9	101	0.00
91	Dibenzo(a,h)anthracene	40.000	39.300	1.8	101	0.00
92	Benzo(g,h,i)perylene	40.000	38.306	4.2	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021521\
Data File : BF123202.D
Acq On : 15 Feb 2021 13:16
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 15 14:00:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 12 16:59:49 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0