

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021522\
 Data File : BF126884.D
 Acq On : 15 Feb 2022 11:22
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 16 00:14:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	0.000	100.0#	0	-0.02
3	Pyridine	40.000	39.098	2.3	94	0.00
4	n-Nitrosodimethylamine	40.000	39.009	2.5	94	0.00
5 S	2-Fluorophenol	80.000	78.594	1.8	96	0.00
6	Aniline	40.000	39.067	2.3	94	0.00
7 S	Phenol-d6	80.000	77.884	2.6	94	0.00
8	2-Chlorophenol	40.000	39.402	1.5	95	0.00
9	Benzaldehyde	40.000	36.118	9.7	93	0.00
10 C	Phenol	40.000	39.038	2.4	95	0.00
11	bis(2-Chloroethyl)ether	40.000	38.239	4.4	92	0.00
12	1,3-Dichlorobenzene	40.000	38.969	2.6	94	0.00
13 C	1,4-Dichlorobenzene	40.000	39.336	1.7	95	0.00
14	1,2-Dichlorobenzene	40.000	38.716	3.2	94	0.00
15	Benzyl Alcohol	40.000	41.414	-3.5	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.784	8.0	88	0.00
17	2-Methylphenol	40.000	38.916	2.7	93	0.00
18	Hexachloroethane	40.000	39.472	1.3	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.525	3.7	94	0.00
20	3+4-Methylphenols	40.000	39.511	1.2	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	38.480	3.8	95	0.00
23 S	Nitrobenzene-d5	80.000	77.360	3.3	94	0.00
24	Nitrobenzene	40.000	39.012	2.5	94	0.00
25	Isophorone	40.000	37.824	5.4	92	0.00
26 C	2-Nitrophenol	40.000	41.333	-3.3	97	0.00
27	2,4-Dimethylphenol	40.000	38.698	3.3	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.092	7.3	90	0.00
29 C	2,4-Dichlorophenol	40.000	38.459	3.9	94	0.00
30	1,2,4-Trichlorobenzene	40.000	38.512	3.7	94	0.00
31	Naphthalene	40.000	38.314	4.2	94	0.00
32	Benzoic acid	40.000	40.585	-1.5	90	0.00
33	4-Chloroaniline	40.000	38.161	4.6	92	0.00
34 C	Hexachlorobutadiene	40.000	39.265	1.8	98	0.00
35	Caprolactam	40.000	38.953	2.6	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.695	3.3	93	0.00
37	2-Methylnaphthalene	40.000	37.868	5.3	93	0.00
38	1-Methylnaphthalene	40.000	38.057	4.9	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.393	4.0	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.169	4.6	87	0.00
42 S	2,4,6-Tribromophenol	80.000	81.934	-2.4	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.754	0.6	96	0.00
44	2,4,5-Trichlorophenol	40.000	39.416	1.5	94	0.00
45 S	2-Fluorobiphenyl	80.000	77.311	3.4	96	0.00
46	1,1'-Biphenyl	40.000	38.405	4.0	94	0.00
47	2-Chloronaphthalene	40.000	38.597	3.5	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021522\
 Data File : BF126884.D
 Acq On : 15 Feb 2022 11:22
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 16 00:14:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 10 00:44:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.423	-1.1	94	0.00
49	Acenaphthylene	40.000	38.381	4.0	92	0.00
50	Dimethylphthalate	40.000	38.000	5.0	93	0.00
51	2,6-Dinitrotoluene	40.000	39.863	0.3	93	0.00
52 C	Acenaphthene	40.000	38.807	3.0	93	0.00
53	3-Nitroaniline	40.000	39.467	1.3	93	0.00
54 P	2,4-Dinitrophenol	40.000	37.792	5.5	89	0.00
55	Dibenzofuran	40.000	38.586	3.5	94	0.00
56 P	4-Nitrophenol	40.000	40.555	-1.4	91	0.00
57	2,4-Dinitrotoluene	40.000	40.311	-0.8	94	0.00
58	Fluorene	40.000	38.858	2.9	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.354	-0.9	97	0.00
60	Diethylphthalate	40.000	38.162	4.6	92	0.00
61	4-Chlorophenyl-phenylether	40.000	38.675	3.3	96	0.00
62	4-Nitroaniline	40.000	40.713	-1.8	93	0.00
63	Azobenzene	40.000	38.230	4.4	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.356	-0.9	92	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.271	6.8	92	0.00
67	4-Bromophenyl-phenylether	40.000	38.686	3.3	95	0.00
68	Hexachlorobenzene	40.000	38.598	3.5	94	0.00
69	Atrazine	40.000	39.149	2.1	93	0.00
70 C	Pentachlorophenol	40.000	41.000	-2.5	93	0.00
71	Phenanthrene	40.000	37.995	5.0	93	0.00
72	Anthracene	40.000	37.774	5.6	92	0.00
73	Carbazole	40.000	38.488	3.8	94	0.00
74	Di-n-butylphthalate	40.000	37.714	5.7	91	0.00
75 C	Fluoranthene	40.000	38.927	2.7	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzydine	40.000	41.621	-4.1	96	0.00
78	Pyrene	40.000	38.388	4.0	94	0.00
79 S	Terphenyl-d14	80.000	76.680	4.1	97	0.00
80	Butylbenzylphthalate	40.000	38.768	3.1	93	0.00
81	Benzo(a)anthracene	40.000	38.532	3.7	96	0.00
82	3,3'-Dichlorobenzidine	40.000	41.082	-2.7	101	0.00
83	Chrysene	40.000	38.860	2.9	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.067	7.3	92	0.00
85 c	Di-n-octyl phthalate	40.000	36.844	7.9	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.219	-0.5	95	0.00
88	Benzo(b)fluoranthene	40.000	38.667	3.3	97	0.00
89	Benzo(k)fluoranthene	40.000	39.371	1.6	99	0.00
90 C	Benzo(a)pyrene	40.000	39.810	0.5	98	0.00
91	Dibenzo(a,h)anthracene	40.000	40.401	-1.0	97	0.00
92	Benzo(g,h,i)perylene	40.000	39.585	1.0	94	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021522\
Data File : BF126884.D
Acq On : 15 Feb 2022 11:22
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 16 00:14:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF020922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 10 00:44:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0