

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF021618\
 Data File : BF103051.D
 Acq On : 16 Feb 2018 14:46
 Operator : SJ/JU
 Sample : J1548-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 22A-2

Manual Integrations
 APPROVED

Sohil
 2/19/2018 11:09:28 AM

Quant Time: Feb 16 16:35:58 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 14:34:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	187744	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	762453	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	309164	20.00	ng	0.00
63) Phenanthrene-d10	11.40	188	455621	20.00	ng	0.00
75) Chrysene-d12	14.04	240	327317	20.00	ng	-0.01
86) Perylene-d12	15.54	264	279653	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	982090	79.44	ng	0.00
7) Phenol-d6	6.51	99	1362112	91.51	ng	0.00
23) Nitrobenzene-d5	7.44	82	856318	77.91	ng	0.00
41) 2,4,6-Tribromophenol	10.70	330	109200	43.88	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	1301902	69.88	ng	0.00
78) Terphenyl-d14	12.99	244	926094	66.99	ng	0.00
Target Compounds						
10) Phenol	6.52	94	61144	3.833	ng	98
49) Dimethylphthalate	9.62	163	112545	4.669	ng	99
73) Di-n-butylphthalate	11.96	149	1813416	59.042	ng	99
79) Butylbenzylphthalate	13.46	149	71994	6.764	ng	96
83) Bis(2-ethylhexyl)phthalate	14.02	149	333604	21.217	ng	100
87) Benzo(b)fluoranthene	15.10	252	41684m	2.299	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA_F\DATA\BF021618\
Data File : BF103051.D
Acq On : 16 Feb 2018 14:46
Operator : SJ/JU
Sample : J1548-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
22A-2

Manual Integrations
APPROVED

Sohil
2/19/2018 11:09:28 AM

Quant Time: Feb 16 16:35:58 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF020318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 16 14:34:49 2018
Response via : Initial Calibration

